

Data Path : Z:\VOASRV\HPCHEM1\MSVOA N\DATA\VN111920\
 Data File : VN064770.D
 Acq On : 20 Nov 2020 9:07
 Operator : JC/MD
 Sample : L4779-05 50X
 Misc : 5.00mL/MSVOA N/WATER
 ALS Vial : 52 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 1555

Quant Time: Nov 24 08:30:26 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N111820W.M
 Quant Title : SW846 8260
 QLast Update : Wed Nov 18 15:29:54 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.63	168	216476	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.55	114	444149	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.38	117	488684	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.32	152	222484	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	7.99	65	215839	56.68	ug/l	0.00
Spiked Amount 50.000			Recovery	=	113.36%	
35) Dibromofluoromethane	7.55	113	153008	60.53	ug/l	0.00
Spiked Amount 50.000			Recovery	=	121.06%	
50) Toluene-d8	10.06	98	590522	51.40	ug/l	0.00
Spiked Amount 50.000			Recovery	=	102.80%	
62) 4-Bromofluorobenzene	12.37	95	258272	60.14	ug/l	0.00
Spiked Amount 50.000			Recovery	=	120.28%	
Target Compounds						
52) Toluene	10.13	92	3445	0.439	ug/l	Qvalue 100
84) 1,2,4-Trimethylbenzene	13.01	105	13024	0.973	ug/l	90

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\VOASRV\HPCHEM1\MSVOA N\DATA\VN111920\
Data File : VN064770.D
Acq On : 20 Nov 2020 9:07
Operator : JC/MD
Sample : L4779-05 50X
Misc : 5.00mL/MSVOA N/WATER
ALS Vial : 52 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampled :
1555

Quant Time: Nov 24 08:30:26 2020
Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N111820W.M
Quant Title : SW846 8260
QLast Update : Wed Nov 18 15:29:54 2020
Response via : Initial Calibration

