

Data Path : Z:\VOASRV\HPCHEM1\MSVOA N\DATA\VN111920\
 Data File : VN064732.D
 Acq On : 19 Nov 2020 14:49
 Operator : JC/MD
 Sample : L4783-19
 Misc : 5.00mL/MSVOA N/WATER
 ALS Vial : 11 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 MW-1D

Integration Parameters: RTEINT.P

Integrator: RTE

Smoothing : ON

Sampling : 1

Start Thrs: 0.2

Stop Thrs : 0

Filtering: 5

Min Area: 3 % of largest Peak

Max Peaks: 100

Peak Location: TOP

If leading or trailing edge < 100 prefer < Baseline drop else tangent >

Peak separation: 5

Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N111820W.M

Title : SW846 8260

Signal : TIC

peak #	R.T. min	first scan	max scan	last scan	PK TY	peak height	corr. area	corr. % max.	% of total
1	4.740	906	933	964	rBV2	331466	1176192	7.95%	4.755%
2	4.994	978	1012	1063	rBV	3993717	14803932	100.00%	59.843%
3	7.550	1789	1807	1818	rBV2	177448	456178	3.08%	1.844%
4	7.627	1818	1831	1855	rVB	307449	743976	5.03%	3.007%
5	7.987	1925	1943	1962	rBV	235218	558114	3.77%	2.256%
6	8.097	1962	1977	1995	rVV	100748	238674	1.61%	0.965%
7	8.550	2101	2118	2137	rBV2	543722	1131740	7.64%	4.575%
8	10.058	2572	2587	2601	rBV	859709	1563986	10.56%	6.322%
9	11.379	2984	2998	3019	rBV	877543	1523354	10.29%	6.158%
10	12.376	3290	3308	3323	rBV	688608	1149369	7.76%	4.646%
11	13.315	3587	3600	3623	rBV	832614	1392556	9.41%	5.629%


Sum of corrected areas: 24738071

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA N\DATA\VN111920\
Data File : VN064732.D
Acq On : 19 Nov 2020 14:49
Operator : JC/MD
Sample : L4783-19
Misc : 5.00mL/MSVOA N/WATER
ALS Vial : 11 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
MW-1D

Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_N\METHODS\82N111820W.M
Quant Title : SW846 8260

TIC Library : C:\DATABASE\NIST11.L
TIC Integration Parameters: LSCINT.P



Data Path : Z:\V0ASRV\HPCHEM1\MSVOA N\DATA\VN111920\
Data File : VN064732.D
Acq On : 19 Nov 2020 14:49
Operator : JC/MD
Sample : L4783-19
Misc : 5.00mL/MSVOA N/WATER
ALS Vial : 11 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
MW-1D










Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_N\METHODS\82N111820W.M
Quant Title : SW846 8260

TIC Library : C:\DATABASE\NIST11.L
TIC Integration Parameters: LSCINT.P

Peak Number 1 Amylene hydrate Concentration Rank 1

R.T.	EstConc	Area	Relative to ISTD	R.T.
8.10	10.54 ug/l	238674	1,4-Difluorobenzene	8.55

Hit#	of	5	Tentative ID	MW	MolForm	CAS#	Qual
1			Amylene hydrate	88	C5H12O	000075-85-4	83
2			3-Pentanol, 2-methyl-	102	C6H14O	000565-67-3	45
3			Propanoic acid, 2-hydroxy-2-methyl-	104	C4H8O3	000594-61-6	38
4			1-Hepten-4-ol	114	C7H14O	003521-91-3	12
5			3-Pentanol, 2,4-dimethyl-	116	C7H16O	000600-36-2	9



Data Path : Z:\VOASRV\HPCHEM1\MSVOA_N\DATA\VN111920\
Data File : VN064732.D
Acq On : 19 Nov 2020 14:49
Operator : JC/MD
Sample : L4783-19
Misc : 5.00mL/MSVOA_N/WATER
ALS Vial : 11 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
MW-1D

Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N111820W.M
Quant Title : SW846 8260

TIC Library : C:\DATABASE\NIST11.L
TIC Integration Parameters: LSCINT.P

TIC Top Hit name	RT	EstConc	Units	Response	--Internal Standard--			
					#	RT	Resp	Conc
Amylene hydrate	8.10	10.5	ug/l	238674	2	8.55	1131740	50.0