

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN111921\
 Data File : VN069587.D
 Acq On : 19 Nov 2021 15:53
 Operator : JC/MD
 Sample : M4739-05DL 5X
 Misc : 5.00mL/MSVOA_N/WATER
 ALS Vial : 8 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 MW-23ADL

Quant Time: Nov 19 16:12:46 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N110921W.M
 Quant Title : SW846 8260
 QLast Update : Tue Nov 09 18:18:22 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.091	168	1329542	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.971	114	2307373	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.747	117	2324514	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.675	152	873147	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.440	65	979078	50.788	ug/l	0.00
Spiked Amount 50.000	Range 61 - 141		Recovery = 101.580%			
35) Dibromofluoromethane	8.027	113	664675	48.030	ug/l	0.00
Spiked Amount 50.000	Range 69 - 133		Recovery = 96.060%			
50) Toluene-d8	10.443	98	2928284	53.802	ug/l	0.00
Spiked Amount 50.000	Range 65 - 126		Recovery = 107.600%			
62) 4-Bromofluorobenzene	12.731	95	1038701	51.949	ug/l	0.00
Spiked Amount 50.000	Range 58 - 135		Recovery = 103.900%			
Target Compounds						Qvalue
19) Methyl tert-butyl Ether	5.624	73	2393331	52.480	ug/l	97
22) Diisopropyl ether	6.503	45	61354	1.196	ug/l	93

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN111921\
Data File : VN069587.D
Acq On : 19 Nov 2021 15:53
Operator : JC/MD
Sample : M4739-05DL 5X
Misc : 5.00mL/MSVOA_N/WATER
ALS Vial : 8 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
MW-23ADL

Quant Time: Nov 19 16:12:46 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N110921W.M
Quant Title : SW846 8260
QLast Update : Tue Nov 09 18:18:22 2021
Response via : Initial Calibration

