

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN111921\
 Data File : VN069589.D
 Acq On : 19 Nov 2021 16:57
 Operator : JC/MD
 Sample : M4764-06
 Misc : 5.00mL/MSVOA_N/WATER
 ALS Vial : 9 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 TRIP

Quant Time: Nov 22 05:08:21 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N110921W.M
 Quant Title : SW846 8260
 QLast Update : Tue Nov 09 18:18:22 2021
 Response via : Initial Calibration

Compound	R.T. QIon		Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.088	168	1596443	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.968	114	2747864	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.747	117	2686730	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.678	152	967656	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.440	65	1165950	50.370	ug/l	0.00
Spiked Amount	50.000	Range 61 - 141	Recovery	= 100.740%		
35) Dibromofluoromethane	8.024	113	804218	48.798	ug/l	0.00
Spiked Amount	50.000	Range 69 - 133	Recovery	= 97.600%		
50) Toluene-d8	10.443	98	3457062	53.335	ug/l	0.00
Spiked Amount	50.000	Range 65 - 126	Recovery	= 106.680%		
62) 4-Bromofluorobenzene	12.731	95	1199489	50.374	ug/l	0.00
Spiked Amount	50.000	Range 58 - 135	Recovery	= 100.740%		
Target Compounds						
						Qvalue
20) Methylene Chloride	5.106	84	10778	0.609	ug/l	93
65) Chlorobenzene	11.771	112	15846	0.317	ug/l	96
95) Naphthalene	15.504	128	9310	4.649	ug/l #	93

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN111921\
Data File : VN069589.D
Acq On : 19 Nov 2021 16:57
Operator : JC/MD
Sample : M4764-06
Misc : 5.00mL/MSVOA_N/WATER
ALS Vial : 9 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
TRIP

Quant Time: Nov 22 05:08:21 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N110921W.M
Quant Title : SW846 8260
QLast Update : Tue Nov 09 18:18:22 2021
Response via : Initial Calibration

