

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN111921\
 Data File : VN069590.D
 Acq On : 19 Nov 2021 17:22
 Operator : JC/MD
 Sample : M4768-04
 Misc : 5.00mL/MSVOA_N/WATER
 ALS Vial : 10 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 VO-TRIP

Quant Time: Nov 22 05:08:31 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N110921W.M
 Quant Title : SW846 8260
 QLast Update : Tue Nov 09 18:18:22 2021
 Response via : Initial Calibration

Compound	R.T. QIon		Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.088	168	1446563	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.968	114	2495116	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.747	117	2416008	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.675	152	859392	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.440	65	1066040	50.826	ug/l	0.00
Spiked Amount	50.000	Range	61 - 141	Recovery	=	101.660%
35) Dibromofluoromethane	8.024	113	735970	49.181	ug/l	0.00
Spiked Amount	50.000	Range	69 - 133	Recovery	=	98.360%
50) Toluene-d8	10.443	98	3120793	53.025	ug/l	0.00
Spiked Amount	50.000	Range	65 - 126	Recovery	=	106.040%
62) 4-Bromofluorobenzene	12.733	95	1064193	49.219	ug/l	0.00
Spiked Amount	50.000	Range	58 - 135	Recovery	=	98.440%
Target Compounds						Qvalue
16) Acetone	4.299	43	24505	2.165	ug/l	# 89

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN111921\
Data File : VN069590.D
Acq On : 19 Nov 2021 17:22
Operator : JC/MD
Sample : M4768-04
Misc : 5.00mL/MSVOA_N/WATER
ALS Vial : 10 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
VO-TRIP

Quant Time: Nov 22 05:08:31 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N110921W.M
Quant Title : SW846 8260
QLast Update : Tue Nov 09 18:18:22 2021
Response via : Initial Calibration

