

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN111921\
 Data File : VN069591.D
 Acq On : 19 Nov 2021 17:48
 Operator : JC/MD
 Sample : M4768-05
 Misc : 5.00mL/MSVOA_N/WATER
 ALS Vial : 11 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 VO-FIELD

Quant Time: Nov 22 05:08:40 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N110921W.M
 Quant Title : SW846 8260
 QLast Update : Tue Nov 09 18:18:22 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.088	168	1315876	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.968	114	2292248	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.746	117	2228848	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.675	152	780067	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.440	65	974774	51.090	ug/l	0.00
Spiked Amount	50.000	Range	61 - 141	Recovery	=	102.180%
35) Dibromofluoromethane	8.024	113	678179	49.330	ug/l	0.00
Spiked Amount	50.000	Range	69 - 133	Recovery	=	98.660%
50) Toluene-d8	10.443	98	2886660	53.387	ug/l	0.00
Spiked Amount	50.000	Range	65 - 126	Recovery	=	106.780%
62) 4-Bromofluorobenzene	12.731	95	984204	49.549	ug/l	0.00
Spiked Amount	50.000	Range	58 - 135	Recovery	=	99.100%
Target Compounds						
16) Acetone	4.296	43	40124	3.898	ug/l	Qvalue 96

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN111921\
Data File : VN069591.D
Acq On : 19 Nov 2021 17:48
Operator : JC/MD
Sample : M4768-05
Misc : 5.00mL/MSVOA_N/WATER
ALS Vial : 11 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
VO-FIELD

Quant Time: Nov 22 05:08:40 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N110921W.M
Quant Title : SW846 8260
QLast Update : Tue Nov 09 18:18:22 2021
Response via : Initial Calibration

