

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN111921\
 Data File : VN069595.D
 Acq On : 19 Nov 2021 19:30
 Operator : JC/MD
 Sample : M4764-04
 Misc : 5.00mL/MSVOA_N/WATER
 ALS Vial : 15 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 MW-6

Quant Time: Nov 22 05:09:25 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N110921W.M
 Quant Title : SW846 8260
 QLast Update : Tue Nov 09 18:18:22 2021
 Response via : Initial Calibration

Compound	R.T. QIon		Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.088	168	1410684	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.971	114	2446169	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.746	117	2432701	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.678	152	896910	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.440	65	1028905	50.303	ug/l	0.00
Spiked Amount 50.000	Range 61 - 141		Recovery	= 100.600%		
35) Dibromofluoromethane	8.024	113	716455	48.834	ug/l	0.00
Spiked Amount 50.000	Range 69 - 133		Recovery	= 97.660%		
50) Toluene-d8	10.443	98	3082798	53.427	ug/l	0.00
Spiked Amount 50.000	Range 65 - 126		Recovery	= 106.860%		
62) 4-Bromofluorobenzene	12.733	95	1104635	52.112	ug/l	0.00
Spiked Amount 50.000	Range 58 - 135		Recovery	= 104.220%		
Target Compounds						
						Qvalue
7) Trichlorofluoromethane	3.379	101	35804	1.324	ug/l	90
27) cis-1,2-Dichloroethene	7.329	96	12552	0.793	ug/l	90
44) Trichloroethene	9.217	130	96628	6.355	ug/l	93
64) Tetrachloroethene	10.979	164	407008	27.708	ug/l	95
65) Chlorobenzene	11.768	112	12334	0.272	ug/l	94

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN111921\
Data File : VN069595.D
Acq On : 19 Nov 2021 19:30
Operator : JC/MD
Sample : M4764-04
Misc : 5.00mL/MSVOA_N/WATER
ALS Vial : 15 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
MW-6

Quant Time: Nov 22 05:09:25 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N110921W.M
Quant Title : SW846 8260
QLast Update : Tue Nov 09 18:18:22 2021
Response via : Initial Calibration

