

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN111924\
 Data File : VN084941.D
 Acq On : 19 Nov 2024 12:52
 Operator : JC\MD
 Sample : P4770-02 500X
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 8 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 MW-3

Quant Time: Nov 20 00:26:42 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N103024W.M
 Quant Title : SW846 8260
 QLast Update : Thu Oct 31 18:45:38 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.218	168	194823	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	9.094	114	349222	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.865	117	306335	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.788	152	148014	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.577	65	140718	50.008	ug/l	0.00
Spiked Amount	50.000	Range 74 - 125	Recovery	= 100.020%		
35) Dibromofluoromethane	8.165	113	113192	47.885	ug/l	0.00
Spiked Amount	50.000	Range 75 - 124	Recovery	= 95.780%		
50) Toluene-d8	10.565	98	413552	47.494	ug/l	0.00
Spiked Amount	50.000	Range 86 - 113	Recovery	= 94.980%		
62) 4-Bromofluorobenzene	12.847	95	163672	50.294	ug/l	0.00
Spiked Amount	50.000	Range 77 - 121	Recovery	= 100.580%		
Target Compounds						
					Qvalue	
39) Methylcyclohexane	9.600	83	2800	0.840	ug/l #	68
40) Benzene	8.600	78	8506	0.808	ug/l	91
52) Toluene	10.630	92	312079	50.684	ug/l	100
67) Ethyl Benzene	11.959	91	130493	11.407	ug/l	99
68) m/p-Xylenes	12.065	106	157125	36.280	ug/l	99
69) o-Xylene	12.394	106	53497	13.204	ug/l	100
73) Isopropylbenzene	12.694	105	31866	3.072	ug/l	98
78) n-propylbenzene	13.029	91	50739	4.174	ug/l	100
80) 1,3,5-Trimethylbenzene	13.171	105	57924	6.738	ug/l	100
84) 1,2,4-Trimethylbenzene	13.482	105	190785	21.502	ug/l	99
95) Naphthalene	15.641	128	36609	4.273	ug/l	99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN111924\
Data File : VN084941.D
Acq On : 19 Nov 2024 12:52
Operator : JC\MD
Sample : P4770-02 500X
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 8 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
MW-3

Quant Time: Nov 20 00:26:42 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N103024W.M
Quant Title : SW846 8260
QLast Update : Thu Oct 31 18:45:38 2024
Response via : Initial Calibration

