

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN111924\
 Data File : VN084943.D
 Acq On : 19 Nov 2024 13:41
 Operator : JC\MD
 Sample : P4845-04DL 5X
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 10 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 FSND-MW-22R-20241112DL

Quant Time: Nov 20 00:28:05 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N103024W.M
 Quant Title : SW846 8260
 QLast Update : Thu Oct 31 18:45:38 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.224	168	185321	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	9.100	114	330213	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.865	117	288531	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.788	152	124157	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.577	65	136208	50.887	ug/l	0.00
Spiked Amount	50.000	Range 74 - 125	Recovery	=	101.780%	
35) Dibromofluoromethane	8.165	113	107837	48.246	ug/l	0.00
Spiked Amount	50.000	Range 75 - 124	Recovery	=	96.500%	
50) Toluene-d8	10.565	98	382722	46.483	ug/l	0.00
Spiked Amount	50.000	Range 86 - 113	Recovery	=	92.960%	
62) 4-Bromofluorobenzene	12.847	95	141731	46.059	ug/l	0.00
Spiked Amount	50.000	Range 77 - 121	Recovery	=	92.120%	
Target Compounds						
44) Trichloroethene	9.347	130	62034	27.020	ug/l	99
64) Tetrachloroethene	11.106	164	1591	0.829	ug/l	91

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN111924\
Data File : VN084943.D
Acq On : 19 Nov 2024 13:41
Operator : JC\MD
Sample : P4845-04DL 5X
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 10 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
FSND-MW-22R-20241112DL

Quant Time: Nov 20 00:28:05 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N103024W.M
Quant Title : SW846 8260
QLast Update : Thu Oct 31 18:45:38 2024
Response via : Initial Calibration

