

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN111924\
 Data File : VN084944.D
 Acq On : 19 Nov 2024 14:05
 Operator : JC\MD
 Sample : P4845-11DL 5X
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 11 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 FSND-MW-33-20241112DL

Quant Time: Nov 20 00:28:39 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N103024W.M
 Quant Title : SW846 8260
 QLast Update : Thu Oct 31 18:45:38 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.224	168	177869	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	9.100	114	318918	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.865	117	284593	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.788	152	120391	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.577	65	133368	51.914	ug/l	0.00
Spiked Amount 50.000	Range 74 - 125		Recovery = 103.820%			
35) Dibromofluoromethane	8.159	113	104519	48.418	ug/l	0.00
Spiked Amount 50.000	Range 75 - 124		Recovery = 96.840%			
50) Toluene-d8	10.565	98	367373	46.199	ug/l	0.00
Spiked Amount 50.000	Range 86 - 113		Recovery = 92.400%			
62) 4-Bromofluorobenzene	12.847	95	139783	47.035	ug/l	0.00
Spiked Amount 50.000	Range 77 - 121		Recovery = 94.080%			
Target Compounds						
44) Trichloroethene	9.347	130	62151	28.029	ug/l	Qvalue 93

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN111924\
Data File : VN084944.D
Acq On : 19 Nov 2024 14:05
Operator : JC\MD
Sample : P4845-11DL 5X
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 11 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
FSND-MW-33-20241112DL

Quant Time: Nov 20 00:28:39 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N103024W.M
Quant Title : SW846 8260
QLast Update : Thu Oct 31 18:45:38 2024
Response via : Initial Calibration

