

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN111924\
 Data File : VN084945.D
 Acq On : 19 Nov 2024 14:29
 Operator : JC\MD
 Sample : P4845-19DL 5X
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 12 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 FSND-GEP-2-20241111DL

Quant Time: Nov 20 00:29:12 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N103024W.M
 Quant Title : SW846 8260
 QLast Update : Thu Oct 31 18:45:38 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.218	168	178600	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	9.100	114	324130	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.865	117	285884	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.794	152	123358	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.577	65	133492	51.749	ug/l	0.00
Spiked Amount 50.000	Range 74 - 125		Recovery = 103.500%			
35) Dibromofluoromethane	8.165	113	104996	47.857	ug/l	0.00
Spiked Amount 50.000	Range 75 - 124		Recovery = 95.720%			
50) Toluene-d8	10.565	98	375860	46.507	ug/l	0.00
Spiked Amount 50.000	Range 86 - 113		Recovery = 93.020%			
62) 4-Bromofluorobenzene	12.847	95	143610	47.546	ug/l	0.00
Spiked Amount 50.000	Range 77 - 121		Recovery = 95.100%			
Target Compounds						
44) Trichloroethene	9.347	130	78042	34.630	ug/l	99
64) Tetrachloroethene	11.100	164	1733	0.911	ug/l #	81

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN111924\
Data File : VN084945.D
Acq On : 19 Nov 2024 14:29
Operator : JC\MD
Sample : P4845-19DL 5X
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 12 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
FSND-GEP-2-20241111DL

Quant Time: Nov 20 00:29:12 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N103024W.M
Quant Title : SW846 8260
QLast Update : Thu Oct 31 18:45:38 2024
Response via : Initial Calibration

