

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN111924\
 Data File : VN084949.D
 Acq On : 19 Nov 2024 16:04
 Operator : JC\MD
 Sample : P4871-01DL 100X
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 16 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 WC-10-A-202411DL

Quant Time: Nov 19 18:32:23 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N103024W.M
 Quant Title : SW846 8260
 QLast Update : Thu Oct 31 18:45:38 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.218	168	165251	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	9.100	114	308690	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.865	117	269378	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.788	152	122161	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.577	65	124265	52.064	ug/l	0.00
Spiked Amount 50.000	Range 74 - 125		Recovery =	104.120%		
35) Dibromofluoromethane	8.165	113	98626	47.202	ug/l	0.00
Spiked Amount 50.000	Range 75 - 124		Recovery =	94.400%		
50) Toluene-d8	10.565	98	353588	45.939	ug/l	0.00
Spiked Amount 50.000	Range 86 - 113		Recovery =	91.880%		
62) 4-Bromofluorobenzene	12.853	95	137811	47.908	ug/l	0.00
Spiked Amount 50.000	Range 77 - 121		Recovery =	95.820%		
Target Compounds						
27) cis-1,2-Dichloroethene	7.477	96	13745	6.092	ug/l	88
44) Trichloroethene	9.359	130	4725	2.202	ug/l	89
52) Toluene	10.635	92	1326	0.244	ug/l #	69
64) Tetrachloroethene	11.100	164	44911	25.065	ug/l	96
68) m/p-Xylenes	12.065	106	3297	0.866	ug/l	98
69) o-Xylene	12.400	106	1536	0.431	ug/l	96
80) 1,3,5-Trimethylbenzene	13.170	105	2030	0.286	ug/l	95
84) 1,2,4-Trimethylbenzene	13.482	105	5373	0.734	ug/l	99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN111924\
Data File : VN084949.D
Acq On : 19 Nov 2024 16:04
Operator : JC\MD
Sample : P4871-01DL 100X
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 16 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
WC-10-A-202411DL

Quant Time: Nov 19 18:32:23 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N103024W.M
Quant Title : SW846 8260
QLast Update : Thu Oct 31 18:45:38 2024
Response via : Initial Calibration

