

Data Path : Z:\VOASRV\HPCHEM1\MSVOA_N\DATA\VN112018\
 Data File : VN052400.D
 Acq On : 20 Nov 2018 11:51
 Operator : MD\SY
 Sample : VN1120WBL01
 Misc : 5.00mL/MSVOA_N/WATER
 ALS Vial : 3 Sample Multiplier: 28

Instrument :
 MSVOA_N
 ClientSampleId :
 VN1120WBL01

Quant Time: Nov 20 13:26:17 2018
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N111418W.M
 Quant Title : SW846 8260
 QLast Update : Thu Nov 15 03:02:37 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.67	168	310856	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.59	114	520250	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.41	117	424174	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.34	152	134230	50.00	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.03	65	230113	52.41	ug/l	0.00
Spiked Amount	50.000		Recovery	=	104.82%	
35) Dibromofluoromethane	7.59	113	170841	50.91	ug/l	0.00
Spiked Amount	50.000		Recovery	=	101.82%	
50) Toluene-d8	10.09	98	647474	49.05	ug/l	0.00
Spiked Amount	50.000		Recovery	=	98.10%	
62) 4-Bromofluorobenzene	12.40	95	182101	41.93	ug/l	0.00
Spiked Amount	50.000		Recovery	=	83.86%	

Target Compounds	Qvalue
------------------	--------

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA_N\DATA\VN112018\
Data File : VN052400.D
Acq On : 20 Nov 2018 11:51
Operator : MD\SY
Sample : VN1120WBL01
Misc : 5.00mL/MSVOA_N/WATER
ALS Vial : 3 Sample Multiplier: 28

Instrument :
MSVOA_N
ClientSampleId :
VN1120WBL01

Quant Time: Nov 20 13:26:17 2018
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_N\METHODS\82N111418W.M
Quant Title : SW846 8260
QLast Update : Thu Nov 15 03:02:37 2018
Response via : Initial Calibration

