

Data Path : Z:\VOASRV\HPCHEM1\MSVOA_N\DATA\VN112018\
 Data File : VN052402.D
 Acq On : 20 Nov 2018 12:52
 Operator : MD\SY
 Sample : PB114954TB
 Misc : 5.00mL/MSVOA_N/WATER
 ALS Vial : 5 Sample Multiplier: 28

Instrument :
 MSVOA_N
 ClientSampleId :
 PB114954TB

Quant Time: Nov 20 13:30:51 2018
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N111418W.M
 Quant Title : SW846 8260
 QLast Update : Thu Nov 15 03:02:37 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.67	168	298609	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.59	114	498862	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.41	117	413121	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.34	152	136748	50.00	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.03	65	218586	51.83	ug/l	0.00
Spiked Amount	50.000		Recovery	=	103.66%	
35) Dibromofluoromethane	7.59	113	164313	51.07	ug/l	0.00
Spiked Amount	50.000		Recovery	=	102.14%	
50) Toluene-d8	10.09	98	622199	49.16	ug/l	0.00
Spiked Amount	50.000		Recovery	=	98.32%	
62) 4-Bromofluorobenzene	12.40	95	178434	42.85	ug/l	0.00
Spiked Amount	50.000		Recovery	=	85.70%	

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA_N\DATA\VN112018\
Data File : VN052402.D
Acq On : 20 Nov 2018 12:52
Operator : MD\SY
Sample : PB114954TB
Misc : 5.00mL/MSVOA_N/WATER
ALS Vial : 5 Sample Multiplier: 28

Instrument :
MSVOA_N
ClientSampleId :
PB114954TB

Quant Time: Nov 20 13:30:51 2018
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_N\METHODS\82N111418W.M
Quant Title : SW846 8260
QLast Update : Thu Nov 15 03:02:37 2018
Response via : Initial Calibration

