

Data Path : Z:\VOASRV\HPCHEM1\MSVOA_N\DATA\VN112018\
 Data File : VN052404.D
 Acq On : 20 Nov 2018 13:45
 Operator : MD\SY
 Sample : J6003-20
 Misc : 5.00mL/MSVOA_N/WATER
 ALS Vial : 7 Sample Multiplier: 28

Instrument :
 MSVOA_N
 ClientSampleId :
 AOC2-03A

Quant Time: Nov 20 14:39:35 2018
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N111418W.M
 Quant Title : SW846 8260
 QLast Update : Thu Nov 15 03:02:37 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.67	168	293763	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.59	114	481731	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.41	117	397736	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.34	152	134215	50.00	ug/l	0.00

System Monitoring Compounds	R.T.	QIon	Response	Conc	Units	Dev(Min)
33) 1,2-Dichloroethane-d4	8.03	65	207126	49.92	ug/l	0.00
Spiked Amount			Recovery	=	99.84%	
35) Dibromofluoromethane	7.59	113	157055	50.55	ug/l	0.00
Spiked Amount			Recovery	=	101.10%	
50) Toluene-d8	10.09	98	598897	49.00	ug/l	0.00
Spiked Amount			Recovery	=	98.00%	
62) 4-Bromofluorobenzene	12.40	95	169590	42.18	ug/l	0.00
Spiked Amount			Recovery	=	84.36%	

Target Compounds	R.T.	QIon	Response	Conc	Units	Qvalue
16) Acetone	3.82	43	7805	6.83	ug/l	96
20) Methylene Chloride	4.55	84	13671	3.28	ug/l	97
25) 2-Butanone	6.84	43	5055	3.14	ug/l #	85
52) Toluene	10.16	92	4072	0.49	ug/l	96
67) Ethyl Benzene	11.51	91	7287	0.50	ug/l	92
68) m/p-Xylenes	11.62	106	3968	0.74	ug/l	91
69) o-Xylene	11.95	106	1259	0.25	ug/l	85
73) Isopropylbenzene	12.25	105	2866	0.27	ug/l	94
78) n-propylbenzene	12.59	91	6745	0.53	ug/l	99
80) 1,3,5-Trimethylbenzene	12.73	105	4289	0.50	ug/l	98
84) 1,2,4-Trimethylbenzene	13.04	105	4382	0.51	ug/l	91
85) sec-Butylbenzene	13.17	105	2925	0.28	ug/l #	76
95) Naphthalene	15.13	128	716	2.75	ug/l #	69

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA_N\DATA\VN112018\
Data File : VN052404.D
Acq On : 20 Nov 2018 13:45
Operator : MD\SY
Sample : J6003-20
Misc : 5.00mL/MSVOA_N/WATER
ALS Vial : 7 Sample Multiplier: 28

Instrument :
MSVOA_N
ClientSampleId :
AOC2-03A

Quant Time: Nov 20 14:39:35 2018
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_N\METHODS\82N111418W.M
Quant Title : SW846 8260
QLast Update : Thu Nov 15 03:02:37 2018
Response via : Initial Calibration

