

Data Path : Z:\VOASRV\HPCHEM1\MSVOA_N\DATA\VN112018\
 Data File : VN052417.D
 Acq On : 20 Nov 2018 19:32
 Operator : MD\SY
 Sample : J6010-03
 Misc : 5.00mL/MSVOA_N/WATER
 ALS Vial : 20 Sample Multiplier: 28

Instrument :
 MSVOA_N
 ClientSampleId :
 FS-RT3786

Quant Time: Nov 21 00:33:41 2018
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N111418W.M
 Quant Title : SW846 8260
 QLast Update : Thu Nov 15 03:02:37 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.67	168	302755	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.59	114	510866	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.41	117	436738	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.34	152	145913	50.00	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.03	65	226623	53.00	ug/l	0.00
Spiked Amount	50.000		Recovery	=	106.00%	
35) Dibromofluoromethane	7.59	113	166941	50.66	ug/l	0.00
Spiked Amount	50.000		Recovery	=	101.32%	
50) Toluene-d8	10.09	98	648107	50.00	ug/l	0.00
Spiked Amount	50.000		Recovery	=	100.00%	
62) 4-Bromofluorobenzene	12.40	95	188261	44.15	ug/l	0.00
Spiked Amount	50.000		Recovery	=	88.30%	

Target Compounds						Qvalue
16) Acetone	3.82	43	6591	5.60	ug/l	95
20) Methylene Chloride	4.56	84	619469	174.70	ug/l	98
43) Isopropyl Acetate	8.17	43	20159	2.81	ug/l #	88

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\VOASRV\HPCHEM1\MSVOA_N\DATA\VN112018\
Data File : VN052417.D
Acq On : 20 Nov 2018 19:32
Operator : MD\SY
Sample : J6010-03
Misc : 5.00mL/MSVOA_N/WATER
ALS Vial : 20 Sample Multiplier: 28

Instrument :
MSVOA_N
ClientSampleId :
FS-RT3786

Quant Time: Nov 21 00:33:41 2018
Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N111418W.M
Quant Title : SW846 8260
QLast Update : Thu Nov 15 03:02:37 2018
Response via : Initial Calibration

