

Data Path : Z:\VOASRV\HPCHEM1\MSVOA_N\DATA\VN112018\
 Data File : VN052418.D
 Acq On : 20 Nov 2018 19:59
 Operator : MD\SY
 Sample : J5775-02
 Misc : 5.00mL/MSVOA_N/WATER
 ALS Vial : 21 Sample Multiplier: 28

Instrument :
 MSVOA_N
 ClientSampleId :
 TR-05-111918

Quant Time: Nov 21 00:36:08 2018
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N111418W.M
 Quant Title : SW846 8260
 QLast Update : Thu Nov 15 03:02:37 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.67	168	302503	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.59	114	506710	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.41	117	430495	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.34	152	142314	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.03	65	224726	52.60	ug/l	0.00
Spiked Amount 50.000			Recovery	=	105.20%	
35) Dibromofluoromethane	7.59	113	168450	51.54	ug/l	0.00
Spiked Amount 50.000			Recovery	=	103.08%	
50) Toluene-d8	10.09	98	643310	50.04	ug/l	0.00
Spiked Amount 50.000			Recovery	=	100.08%	
62) 4-Bromofluorobenzene	12.40	95	184892	43.71	ug/l	0.00
Spiked Amount 50.000			Recovery	=	87.42%	
Target Compounds						
16) Acetone	3.82	43	8721	7.41	ug/l	91
20) Methylene Chloride	4.56	84	32568	8.52	ug/l	97
43) Isopropyl Acetate	8.17	43	17746	2.50	ug/l #	95

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA_N\DATA\VN112018\
Data File : VN052418.D
Acq On : 20 Nov 2018 19:59
Operator : MD\SY
Sample : J5775-02
Misc : 5.00mL/MSVOA_N/WATER
ALS Vial : 21 Sample Multiplier: 28

Instrument :
MSVOA_N
ClientSampleId :
TR-05-111918

Quant Time: Nov 21 00:36:08 2018
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_N\METHODS\82N111418W.M
Quant Title : SW846 8260
QLast Update : Thu Nov 15 03:02:37 2018
Response via : Initial Calibration

