

Data Path : Z:\VOASRV\HPCHEM1\MSVOA_N\DATA\VN112018\
 Data File : VN052421.D
 Acq On : 20 Nov 2018 21:18
 Operator : MD\SY
 Sample : J5999-02
 Misc : 5.00mL/MSVOA_N/WATER
 ALS Vial : 24 Sample Multiplier: 28

Instrument :
 MSVOA_N
 ClientSampleId :
 SSSI-1114-S-TCLP-3

Manual Integrations
 APPROVED

MMDadoda
 11/21/2018 5:12:50 PM

Quant Time: Nov 21 00:50:57 2018
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N111418W.M
 Quant Title : SW846 8260
 QLast Update : Thu Nov 15 03:02:37 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.67	168	321336	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.59	114	533866	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.41	117	471931	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.34	152	177809	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.03	65	240373	52.96	ug/l	0.00
Spiked Amount 50.000			Recovery	=	105.92%	
35) Dibromofluoromethane	7.59	113	177485	51.54	ug/l	0.00
Spiked Amount 50.000			Recovery	=	103.08%	
50) Toluene-d8	10.09	98	690160	50.95	ug/l	0.00
Spiked Amount 50.000			Recovery	=	101.90%	
62) 4-Bromofluorobenzene	12.40	95	214805	48.20	ug/l	0.00
Spiked Amount 50.000			Recovery	=	96.40%	
Target Compounds						
16) Acetone	3.82	43	9844	7.88	ug/l	93
20) Methylene Chloride	4.51	84	42559473m	11353.47	ug/l	
40) Benzene	8.04	78	27638	1.81	ug/l	91
43) Isopropyl Acetate	8.17	43	16020	2.14	ug/l #	93
52) Toluene	10.16	92	239021	26.14	ug/l	99
67) Ethyl Benzene	11.51	91	91166	5.27	ug/l	98
68) m/p-Xylenes	11.62	106	113621	17.96	ug/l	97
69) o-Xylene	11.95	106	49163	8.11	ug/l	97
78) n-propylbenzene	12.59	91	23452	1.40	ug/l	99
80) 1,3,5-Trimethylbenzene	12.73	105	33773	2.97	ug/l	96
84) 1,2,4-Trimethylbenzene	13.04	105	131671	11.54	ug/l	98
95) Naphthalene	15.13	128	25361	6.42	ug/l	99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\VOASRV\HPCHEM1\MSVOA_N\DATA\VN112018\
Data File : VN052421.D
Acq On : 20 Nov 2018 21:18
Operator : MD\SY
Sample : J5999-02
Misc : 5.00mL/MSVOA_N/WATER
ALS Vial : 24 Sample Multiplier: 28

Instrument :
MSVOA_N
ClientSampleId :
SSSI-1114-S-TCLP-3

Manual Integrations
APPROVED

MMDadoda
11/21/2018 5:12:50 PM

Quant Time: Nov 21 00:50:57 2018
Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N111418W.M
Quant Title : SW846 8260
QLast Update : Thu Nov 15 03:02:37 2018
Response via : Initial Calibration

