

Data Path : Z:\VOASRV\HPCHEM1\MSVOA_N\DATA\VN112018\
 Data File : VN052423.D
 Acq On : 20 Nov 2018 22:12
 Operator : MD\SY
 Sample : J6000-02
 Misc : 5.00mL/MSVOA_N/WATER
 ALS Vial : 26 Sample Multiplier: 28

Instrument :
 MSVOA_N
 ClientSampleId :
 FS-12072-11929

Quant Time: Nov 21 01:01:15 2018
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N111418W.M
 Quant Title : SW846 8260
 QLast Update : Thu Nov 15 03:02:37 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.67	168	300448	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.59	114	503098	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.41	117	434192	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.34	152	157141	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.03	65	228052	53.74	ug/l	0.00
Spiked Amount 50.000			Recovery	=	107.48%	
35) Dibromofluoromethane	7.59	113	169844	52.34	ug/l	0.00
Spiked Amount 50.000			Recovery	=	104.68%	
50) Toluene-d8	10.09	98	642848	50.36	ug/l	0.00
Spiked Amount 50.000			Recovery	=	100.72%	
62) 4-Bromofluorobenzene	12.40	95	193007	45.96	ug/l	0.00
Spiked Amount 50.000			Recovery	=	91.92%	
Target Compounds						
16) Acetone	3.82	43	5931	5.08	ug/l #	89
20) Methylene Chloride	4.56	84	153493	43.09	ug/l	97
43) Isopropyl Acetate	8.17	43	21830	3.09	ug/l #	94

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA_N\DATA\VN112018\
Data File : VN052423.D
Acq On : 20 Nov 2018 22:12
Operator : MD\SY
Sample : J6000-02
Misc : 5.00mL/MSVOA_N/WATER
ALS Vial : 26 Sample Multiplier: 28

Instrument :
MSVOA_N
ClientSampleId :
FS-12072-11929

Quant Time: Nov 21 01:01:15 2018
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_N\METHODS\82N111418W.M
Quant Title : SW846 8260
QLast Update : Thu Nov 15 03:02:37 2018
Response via : Initial Calibration

