

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN112023\
 Data File : VN080199.D
 Acq On : 20 Nov 2023 11:38
 Operator : JC\MD
 Sample : VN1120WBL02
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 4 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 VN1120WBL02

Quant Time: Nov 21 01:01:06 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N111623W.M
 Quant Title : SW846 8260
 QLast Update : Fri Nov 17 01:19:32 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.224	168	156510	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	9.100	114	277374	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.865	117	297033	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.794	152	133290	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.582	65	145534	56.316	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	=	112.640%
35) Dibromofluoromethane	8.165	113	110121	53.805	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	107.620%
50) Toluene-d8	10.570	98	372379	50.473	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	=	100.940%
62) 4-Bromofluorobenzene	12.853	95	146461	52.557	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	105.120%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN112023\
Data File : VN080199.D
Acq On : 20 Nov 2023 11:38
Operator : JC\MD
Sample : VN1120WBL02
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 4 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
VN1120WBL02

Quant Time: Nov 21 01:01:06 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N111623W.M
Quant Title : SW846 8260
QLast Update : Fri Nov 17 01:19:32 2023
Response via : Initial Calibration

