

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN112023\
 Data File : VN080203.D
 Acq On : 20 Nov 2023 13:25
 Operator : JC\MD
 Sample : 05481-02 100X
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 8 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 COMP-AQ

Quant Time: Nov 21 01:03:46 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N111623W.M
 Quant Title : SW846 8260
 QLast Update : Fri Nov 17 01:19:32 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.229	168	145997	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	9.106	114	268674	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.865	117	282756	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.794	152	129176	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.582	65	137391	56.994	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	=	113.980%
35) Dibromofluoromethane	8.165	113	104236	52.579	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	105.160%
50) Toluene-d8	10.570	98	355217	49.706	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	=	99.420%
62) 4-Bromofluorobenzene	12.853	95	137135	50.804	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	101.600%
Target Compounds						
						Qvalue
20) Methylene Chloride	5.277	84	3283	1.936	ug/l	92

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN112023\
Data File : VN080203.D
Acq On : 20 Nov 2023 13:25
Operator : JC\MD
Sample : 05481-02 100X
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 8 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
COMP-AQ

Quant Time: Nov 21 01:03:46 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N111623W.M
Quant Title : SW846 8260
QLast Update : Fri Nov 17 01:19:32 2023
Response via : Initial Calibration

