

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN112023\
 Data File : VN080204.D
 Acq On : 20 Nov 2023 13:49
 Operator : JC\MD
 Sample : 05484-02 100X
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 9 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 AUD-1293

Quant Time: Nov 21 01:04:12 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N111623W.M
 Quant Title : SW846 8260
 QLast Update : Fri Nov 17 01:19:32 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.230	168	151603	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	9.106	114	270934	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.865	117	294076	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.794	152	141475	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.577	65	137574	54.959	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	= 109.920%	
35) Dibromofluoromethane	8.165	113	105151	52.598	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	= 105.200%	
50) Toluene-d8	10.571	98	363070	50.381	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	= 100.760%	
62) 4-Bromofluorobenzene	12.853	95	147811	54.302	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	= 108.600%	
Target Compounds						
					Qvalue	
40) Benzene	8.618	78	3563	0.433	ug/l	100
52) Toluene	10.635	92	2879	0.559	ug/l	88
68) m/p-Xylenes	12.071	106	2598	0.601	ug/l	84
95) Naphthalene	15.647	128	14288	1.773	ug/l	98

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN112023\
Data File : VN080204.D
Acq On : 20 Nov 2023 13:49
Operator : JC\MD
Sample : 05484-02 100X
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 9 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
AUD-1293

Quant Time: Nov 21 01:04:12 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N111623W.M
Quant Title : SW846 8260
QLast Update : Fri Nov 17 01:19:32 2023
Response via : Initial Calibration

