

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN112023\
 Data File : VN080205.D
 Acq On : 20 Nov 2023 14:12
 Operator : JC\MD
 Sample : PB157260TB
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 10 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 PB157260TB

Quant Time: Nov 21 01:04:38 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N111623W.M
 Quant Title : SW846 8260
 QLast Update : Fri Nov 17 01:19:32 2023
 Response via : Initial Calibration

Compound	R.T. QIon		Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.230	168	145793	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	9.106	114	271421	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.865	117	309825	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.794	152	133073	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.577	65	135105	56.124	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	=	112.240%
35) Dibromofluoromethane	8.171	113	104930	52.394	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	104.780%
50) Toluene-d8	10.571	98	380863	52.755	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	=	105.520%
62) 4-Bromofluorobenzene	12.853	95	153038	56.122	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	112.240%
Target Compounds						Qvalue
16) Acetone	4.430	43	14774	18.272	ug/l	99
20) Methylene Chloride	5.283	84	19703	11.636	ug/l	92
43) Isopropyl Acetate	8.688	43	101792	26.008	ug/l #	87

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN112023\
Data File : VN080205.D
Acq On : 20 Nov 2023 14:12
Operator : JC\MD
Sample : PB157260TB
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 10 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
PB157260TB

Quant Time: Nov 21 01:04:38 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N111623W.M
Quant Title : SW846 8260
QLast Update : Fri Nov 17 01:19:32 2023
Response via : Initial Calibration

