

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN112023\
 Data File : VN080206.D
 Acq On : 20 Nov 2023 14:36
 Operator : JC\MD
 Sample : PB157250TB
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 11 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 PB157250TB

Quant Time: Nov 21 01:05:05 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N111623W.M
 Quant Title : SW846 8260
 QLast Update : Fri Nov 17 01:19:32 2023
 Response via : Initial Calibration

Compound	R.T. QIon		Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.224	168	143103	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	9.106	114	266108	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.871	117	309728	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.794	152	136299	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.582	65	135479	57.337	ug/l	0.00
Spiked Amount	50.000	Range 74 - 125	Recovery	= 114.680%		
35) Dibromofluoromethane	8.171	113	103844	52.886	ug/l	0.00
Spiked Amount	50.000	Range 75 - 124	Recovery	= 105.780%		
50) Toluene-d8	10.571	98	376338	53.169	ug/l	0.00
Spiked Amount	50.000	Range 86 - 113	Recovery	= 106.340%		
62) 4-Bromofluorobenzene	12.853	95	152544	57.057	ug/l	0.00
Spiked Amount	50.000	Range 83 - 123	Recovery	= 114.120%		
Target Compounds						
						Qvalue
16) Acetone	4.436	43	17027	21.455	ug/l #	75
20) Methylene Chloride	5.283	84	17551	10.560	ug/l #	87
43) Isopropyl Acetate	8.688	43	101768	26.521	ug/l #	88

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN112023\
Data File : VN080206.D
Acq On : 20 Nov 2023 14:36
Operator : JC\MD
Sample : PB157250TB
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 11 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
PB157250TB

Quant Time: Nov 21 01:05:05 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N111623W.M
Quant Title : SW846 8260
QLast Update : Fri Nov 17 01:19:32 2023
Response via : Initial Calibration

