

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN112023\
 Data File : VN080207.D
 Acq On : 20 Nov 2023 15:00
 Operator : JC\MD
 Sample : PB157250ZHE#01
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 12 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 PB157250ZHE#01

Manual Integrations
 APPROVED

Reviewed By : John Carlone 11/21/2023
 Supervised By : Mahesh Dadoda 11/21/2023

Quant Time: Nov 21 01:05:32 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N111623W.M
 Quant Title : SW846 8260
 QLast Update : Fri Nov 17 01:19:32 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.230	168	147917	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	9.106	114	271875	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.865	117	306792	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.794	152	134796	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.577	65	138622	56.758	ug/l	0.00
Spiked Amount 50.000	Range 74 - 125		Recovery = 113.520%			
35) Dibromofluoromethane	8.171	113	107143	53.409	ug/l	0.00
Spiked Amount 50.000	Range 75 - 124		Recovery = 106.820%			
50) Toluene-d8	10.571	98	374319	51.762	ug/l	0.00
Spiked Amount 50.000	Range 86 - 113		Recovery = 103.520%			
62) 4-Bromofluorobenzene	12.847	95	151505	55.467	ug/l	0.00
Spiked Amount 50.000	Range 83 - 123		Recovery = 110.940%			
Target Compounds					Qvalue	
16) Acetone	4.442	43	12534	15.279	ug/l	95
20) Methylene Chloride	5.271	84	14191m	8.261	ug/l	
43) Isopropyl Acetate	8.688	43	97775	24.940	ug/l #	86

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN112023\
Data File : VN080207.D
Acq On : 20 Nov 2023 15:00
Operator : JC\MD
Sample : PB157250ZHE#01
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 12 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
PB157250ZHE#01

Quant Time: Nov 21 01:05:32 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N111623W.M
Quant Title : SW846 8260
QLast Update : Fri Nov 17 01:19:32 2023
Response via : Initial Calibration

Manual Integrations
APPROVED

Reviewed By :John Carlone 11/21/2023
Supervised By :Mahesh Dadoda 11/21/2023

