

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN112023\
 Data File : VN080208.D
 Acq On : 20 Nov 2023 15:24
 Operator : JC\MD
 Sample : PB157250ZHE#02
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 13 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 PB157250ZHE#02

Quant Time: Nov 21 01:05:59 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N111623W.M
 Quant Title : SW846 8260
 QLast Update : Fri Nov 17 01:19:32 2023
 Response via : Initial Calibration

Compound	R.T. QIon		Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.230	168	138000	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	9.106	114	258720	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.865	117	293338	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.794	152	128254	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.582	65	131007	57.495	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	=	114.980%
35) Dibromofluoromethane	8.171	113	99855	52.307	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	104.620%
50) Toluene-d8	10.571	98	360438	52.377	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	=	104.760%
62) 4-Bromofluorobenzene	12.853	95	142635	54.874	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	109.740%
Target Compounds						
						Qvalue
16) Acetone	4.436	43	12978	16.957	ug/l	100
20) Methylene Chloride	5.283	84	14464	9.025	ug/l #	82
43) Isopropyl Acetate	8.688	43	93866	25.161	ug/l #	86

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN112023\
Data File : VN080208.D
Acq On : 20 Nov 2023 15:24
Operator : JC\MD
Sample : PB157250ZHE#02
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 13 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
PB157250ZHE#02

Quant Time: Nov 21 01:05:59 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N111623W.M
Quant Title : SW846 8260
QLast Update : Fri Nov 17 01:19:32 2023
Response via : Initial Calibration

