

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN112023\
 Data File : VN080209.D
 Acq On : 20 Nov 2023 15:48
 Operator : JC\MD
 Sample : PB157250ZHE#03
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 14 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 PB157250ZHE#03

Quant Time: Nov 21 01:06:28 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N111623W.M
 Quant Title : SW846 8260
 QLast Update : Fri Nov 17 01:19:32 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.224	168	144198	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	9.106	114	264877	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.865	117	298295	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.794	152	125354	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.577	65	138161	58.028	ug/l	0.00
Spiked Amount 50.000	Range 74 - 125		Recovery = 116.060%			
35) Dibromofluoromethane	8.165	113	102332	52.359	ug/l	0.00
Spiked Amount 50.000	Range 75 - 124		Recovery = 104.720%			
50) Toluene-d8	10.571	98	372092	52.814	ug/l	0.00
Spiked Amount 50.000	Range 86 - 113		Recovery = 105.620%			
62) 4-Bromofluorobenzene	12.853	95	145132	54.537	ug/l	0.00
Spiked Amount 50.000	Range 83 - 123		Recovery = 109.080%			
Target Compounds						Qvalue
16) Acetone	4.430	43	12589	15.742	ug/l	97
20) Methylene Chloride	5.277	84	17129	10.228	ug/l	89
43) Isopropyl Acetate	8.694	43	99526	26.058	ug/l #	86

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN112023\
Data File : VN080209.D
Acq On : 20 Nov 2023 15:48
Operator : JC\MD
Sample : PB157250ZHE#03
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 14 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
PB157250ZHE#03

Quant Time: Nov 21 01:06:28 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N111623W.M
Quant Title : SW846 8260
QLast Update : Fri Nov 17 01:19:32 2023
Response via : Initial Calibration

