

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN112023\
 Data File : VN080210.D
 Acq On : 20 Nov 2023 16:11
 Operator : JC\MD
 Sample : PB157250ZHE#04
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 15 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 PB157250ZHE#04

Quant Time: Nov 21 01:06:56 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N111623W.M
 Quant Title : SW846 8260
 QLast Update : Fri Nov 17 01:19:32 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.229	168	140235	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	9.106	114	265007	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.870	117	297511	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.794	152	126168	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.582	65	134828	58.228	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	=	116.460%
35) Dibromofluoromethane	8.165	113	100883	51.592	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	103.180%
50) Toluene-d8	10.570	98	368453	52.271	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	=	104.540%
62) 4-Bromofluorobenzene	12.853	95	147166	55.274	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	110.540%
Target Compounds						
						Qvalue
16) Acetone	4.436	43	14478	18.616	ug/l #	78
20) Methylene Chloride	5.277	84	16356	10.043	ug/l #	90
43) Isopropyl Acetate	8.688	43	103425	27.065	ug/l #	85

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN112023\
Data File : VN080210.D
Acq On : 20 Nov 2023 16:11
Operator : JC\MD
Sample : PB157250ZHE#04
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 15 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
PB157250ZHE#04

Quant Time: Nov 21 01:06:56 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N111623W.M
Quant Title : SW846 8260
QLast Update : Fri Nov 17 01:19:32 2023
Response via : Initial Calibration

