

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN112023\
 Data File : VN080218.D
 Acq On : 20 Nov 2023 19:21
 Operator : JC\MD
 Sample : 05469-04
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 23 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 TP-12

Quant Time: Nov 21 01:10:35 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N111623W.M
 Quant Title : SW846 8260
 QLast Update : Fri Nov 17 01:19:32 2023
 Response via : Initial Calibration

Compound	R.T. QIon		Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.229	168	143311	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	9.100	114	265350	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.865	117	296240	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.794	152	126739	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.582	65	134380	56.789	ug/l	0.00
Spiked Amount	50.000	Range 74 - 125	Recovery	= 113.580%		
35) Dibromofluoromethane	8.165	113	101044	51.607	ug/l	0.00
Spiked Amount	50.000	Range 75 - 124	Recovery	= 103.220%		
50) Toluene-d8	10.570	98	363935	51.564	ug/l	0.00
Spiked Amount	50.000	Range 86 - 113	Recovery	= 103.120%		
62) 4-Bromofluorobenzene	12.853	95	144308	54.131	ug/l	0.00
Spiked Amount	50.000	Range 83 - 123	Recovery	= 108.260%		
Target Compounds						
					Qvalue	
16) Acetone	4.430	43	18106	22.781	ug/l	95
18) Methyl Acetate	5.030	43	3156	1.236	ug/l #	81
20) Methylene Chloride	5.277	84	58173	34.952	ug/l	95
43) Isopropyl Acetate	8.694	43	100789	26.341	ug/l #	83

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN112023\
Data File : VN080218.D
Acq On : 20 Nov 2023 19:21
Operator : JC\MD
Sample : 05469-04
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 23 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
TP-12

Quant Time: Nov 21 01:10:35 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N111623W.M
Quant Title : SW846 8260
QLast Update : Fri Nov 17 01:19:32 2023
Response via : Initial Calibration

