

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN112023\
 Data File : VN080219.D
 Acq On : 20 Nov 2023 19:45
 Operator : JC\MD
 Sample : 05489-02
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 24 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 TRACKS-SP-1-E2

Quant Time: Nov 21 01:11:04 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N111623W.M
 Quant Title : SW846 8260
 QLast Update : Fri Nov 17 01:19:32 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.230	168	130820	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	9.106	114	250967	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.871	117	281426	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.788	152	123332	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.583	65	130307	60.326	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	= 120.660%	
35) Dibromofluoromethane	8.171	113	96035	51.860	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	= 103.720%	
50) Toluene-d8	10.571	98	348485	52.204	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	= 104.400%	
62) 4-Bromofluorobenzene	12.847	95	134274	53.254	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	= 106.500%	
Target Compounds						
					Qvalue	
16) Acetone	4.436	43	19234	26.511	ug/l	93
18) Methyl Acetate	5.030	43	2982	1.279	ug/l #	68
20) Methylene Chloride	5.277	84	68222	44.903	ug/l	97
43) Isopropyl Acetate	8.688	43	84630	23.386	ug/l #	80
95) Naphthalene	15.641	128	10568	1.504	ug/l	97

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN112023\
Data File : VN080219.D
Acq On : 20 Nov 2023 19:45
Operator : JC\MD
Sample : 05489-02
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 24 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
TRACKS-SP-1-E2

Quant Time: Nov 21 01:11:04 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N111623W.M
Quant Title : SW846 8260
QLast Update : Fri Nov 17 01:19:32 2023
Response via : Initial Calibration

