

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN112023\
 Data File : VN080220.D
 Acq On : 20 Nov 2023 20:09
 Operator : JC\MD
 Sample : 05489-05
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 25 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 TRACKS-SP-2-E2

Quant Time: Nov 21 01:11:30 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N111623W.M
 Quant Title : SW846 8260
 QLast Update : Fri Nov 17 01:19:32 2023
 Response via : Initial Calibration

Compound	R.T. QIon		Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.230	168	137499	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	9.106	114	261081	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.865	117	288694	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.794	152	123254	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.582	65	133468	58.788	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	=	117.580%
35) Dibromofluoromethane	8.171	113	101261	52.564	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	105.120%
50) Toluene-d8	10.565	98	359849	51.819	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	=	103.640%
62) 4-Bromofluorobenzene	12.853	95	138924	52.963	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	105.920%
Target Compounds						
						Qvalue
16) Acetone	4.424	43	21801	28.590	ug/l #	85
20) Methylene Chloride	5.283	84	59285	37.125	ug/l	96
43) Isopropyl Acetate	8.688	43	88623	23.540	ug/l #	82

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN112023\
Data File : VN080220.D
Acq On : 20 Nov 2023 20:09
Operator : JC\MD
Sample : 05489-05
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 25 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
TRACKS-SP-2-E2

Quant Time: Nov 21 01:11:30 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N111623W.M
Quant Title : SW846 8260
QLast Update : Fri Nov 17 01:19:32 2023
Response via : Initial Calibration

