

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN112023\
 Data File : VN080222.D
 Acq On : 20 Nov 2023 20:56
 Operator : JC\MD
 Sample : 05492-04
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 27 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 TP-13

Quant Time: Nov 21 01:12:23 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N111623W.M
 Quant Title : SW846 8260
 QLast Update : Fri Nov 17 01:19:32 2023
 Response via : Initial Calibration

Compound	R.T. QIon		Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.230	168	136437	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	9.106	114	254651	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.871	117	292127	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.794	152	127843	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.582	65	132325	58.738	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	= 117.480%	
35) Dibromofluoromethane	8.171	113	101642	54.094	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	= 108.180%	
50) Toluene-d8	10.571	98	361809	53.416	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	= 106.840%	
62) 4-Bromofluorobenzene	12.853	95	145118	56.722	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	= 113.440%	
Target Compounds						
						Qvalue
16) Acetone	4.430	43	15488	20.469	ug/l #	84
20) Methylene Chloride	5.283	84	45656	28.813	ug/l	95
43) Isopropyl Acetate	8.688	43	95011	25.874	ug/l #	83

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN112023\
Data File : VN080222.D
Acq On : 20 Nov 2023 20:56
Operator : JC\MD
Sample : 05492-04
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 27 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
TP-13

Quant Time: Nov 21 01:12:23 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N111623W.M
Quant Title : SW846 8260
QLast Update : Fri Nov 17 01:19:32 2023
Response via : Initial Calibration

