

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN112024\
 Data File : VN084966.D
 Acq On : 20 Nov 2024 14:08
 Operator : JC\MD
 Sample : PB165108TB
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 9 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 PB165108TB

Quant Time: Nov 21 00:37:53 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N103024W.M
 Quant Title : SW846 8260
 QLast Update : Thu Oct 31 18:45:38 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.218	168	174556	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	9.100	114	321916	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.865	117	287968	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.794	152	120018	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.576	65	126349	50.115	ug/l	0.00
Spiked Amount	50.000	Range 74 - 125	Recovery	=	100.240%	
35) Dibromofluoromethane	8.165	113	102758	47.159	ug/l	0.00
Spiked Amount	50.000	Range 75 - 124	Recovery	=	94.320%	
50) Toluene-d8	10.565	98	377162	46.989	ug/l	0.00
Spiked Amount	50.000	Range 86 - 113	Recovery	=	93.980%	
62) 4-Bromofluorobenzene	12.847	95	139122	46.377	ug/l	0.00
Spiked Amount	50.000	Range 77 - 121	Recovery	=	92.760%	
Target Compounds						
16) Acetone	4.436	43	19276	23.888	ug/l	98
20) Methylene Chloride	5.271	84	3124	1.409	ug/l #	69
43) Isopropyl Acetate	8.688	43	139307	27.059	ug/l #	85

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN112024\
Data File : VN084966.D
Acq On : 20 Nov 2024 14:08
Operator : JC\MD
Sample : PB165108TB
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 9 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
PB165108TB

Quant Time: Nov 21 00:37:53 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N103024W.M
Quant Title : SW846 8260
QLast Update : Thu Oct 31 18:45:38 2024
Response via : Initial Calibration

