

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN112024\
 Data File : VN084967.D
 Acq On : 20 Nov 2024 14:33
 Operator : JC\MD
 Sample : PB165122TB
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 10 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 PB165122TB

Quant Time: Nov 21 00:38:25 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N103024W.M
 Quant Title : SW846 8260
 QLast Update : Thu Oct 31 18:45:38 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.224	168	165076	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	9.100	114	299710	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.865	117	271233	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.788	152	111994	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.577	65	122130	51.224	ug/l	0.00
Spiked Amount 50.000	Range 74 - 125		Recovery = 102.440%			
35) Dibromofluoromethane	8.165	113	96695	47.664	ug/l	0.00
Spiked Amount 50.000	Range 75 - 124		Recovery = 95.320%			
50) Toluene-d8	10.565	98	354600	47.451	ug/l	0.00
Spiked Amount 50.000	Range 86 - 113		Recovery = 94.900%			
62) 4-Bromofluorobenzene	12.847	95	133518	47.806	ug/l	0.00
Spiked Amount 50.000	Range 77 - 121		Recovery = 95.620%			
Target Compounds						
16) Acetone	4.436	43	19968	26.442	ug/l #	86
20) Methylene Chloride	5.271	84	2825	1.347	ug/l #	63
43) Isopropyl Acetate	8.688	43	242604	51.472	ug/l #	73

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN112024\
Data File : VN084967.D
Acq On : 20 Nov 2024 14:33
Operator : JC\MD
Sample : PB165122TB
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 10 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
PB165122TB

Quant Time: Nov 21 00:38:25 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N103024W.M
Quant Title : SW846 8260
QLast Update : Thu Oct 31 18:45:38 2024
Response via : Initial Calibration

