

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN112024\
 Data File : VN084972.D
 Acq On : 20 Nov 2024 16:34
 Operator : JC\MD
 Sample : P4893-08
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 15 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 MH-762

Quant Time: Nov 21 00:42:10 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N103024W.M
 Quant Title : SW846 8260
 QLast Update : Thu Oct 31 18:45:38 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.224	168	157220	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	9.100	114	293552	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.865	117	266271	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.794	152	115985	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.576	65	116478	51.294	ug/l	0.00
Spiked Amount	50.000	Range 74 - 125	Recovery	=	102.580%	
35) Dibromofluoromethane	8.165	113	95455	48.040	ug/l	0.00
Spiked Amount	50.000	Range 75 - 124	Recovery	=	96.080%	
50) Toluene-d8	10.565	98	349006	47.682	ug/l	0.00
Spiked Amount	50.000	Range 86 - 113	Recovery	=	95.360%	
62) 4-Bromofluorobenzene	12.847	95	133505	48.804	ug/l	0.00
Spiked Amount	50.000	Range 77 - 121	Recovery	=	97.600%	
Target Compounds						
16) Acetone	4.436	43	20859	29.282	ug/l	97
43) Isopropyl Acetate	8.688	43	177014	38.093	ug/l #	77

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN112024\
Data File : VN084972.D
Acq On : 20 Nov 2024 16:34
Operator : JC\MD
Sample : P4893-08
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 15 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
MH-762

Quant Time: Nov 21 00:42:10 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N103024W.M
Quant Title : SW846 8260
QLast Update : Thu Oct 31 18:45:38 2024
Response via : Initial Calibration

