

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN112024\
 Data File : VN084974.D
 Acq On : 20 Nov 2024 17:22
 Operator : JC\MD
 Sample : P4910-08
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 17 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 MH-759

Manual Integrations
 APPROVED

Reviewed By :Semsettin Yesilyurt 11/21/2024
 Supervised By :Mahesh Dadoda 11/21/2024

Quant Time: Nov 21 00:43:20 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N103024W.M
 Quant Title : SW846 8260
 QLast Update : Thu Oct 31 18:45:38 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.224	168	167654	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	9.100	114	308160	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.865	117	286228	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.794	152	122831	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.576	65	125704	51.912	ug/l	0.00
Spiked Amount 50.000	Range 74 - 125		Recovery =	103.820%		
35) Dibromofluoromethane	8.165	113	100903	48.375	ug/l	0.00
Spiked Amount 50.000	Range 75 - 124		Recovery =	96.740%		
50) Toluene-d8	10.565	98	365667	47.590	ug/l	0.00
Spiked Amount 50.000	Range 86 - 113		Recovery =	95.180%		
62) 4-Bromofluorobenzene	12.847	95	142382	49.582	ug/l	0.00
Spiked Amount 50.000	Range 77 - 121		Recovery =	99.160%		
Target Compounds						
16) Acetone	4.436	43	18948	24.516	ug/l	Qvalue 100
43) Isopropyl Acetate	8.688	43	154543m	31.515	ug/l	

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN112024\
Data File : VN084974.D
Acq On : 20 Nov 2024 17:22
Operator : JC\MD
Sample : P4910-08
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 17 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
MH-759

Manual Integrations
APPROVED

Reviewed By :Semsettin Yesilyurt 11/21/2024
Supervised By :Mahesh Dadoda 11/21/2024

Quant Time: Nov 21 00:43:20 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N103024W.M
Quant Title : SW846 8260
QLast Update : Thu Oct 31 18:45:38 2024
Response via : Initial Calibration

