

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN112024\
 Data File : VN084977.D
 Acq On : 20 Nov 2024 18:34
 Operator : JC\MD
 Sample : P4916-12
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 20 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 TP-3-WC

Manual Integrations
 APPROVED

Reviewed By :Semsettin Yesilyurt 11/21/2024
 Supervised By :Mahesh Dadoda 11/21/2024

Quant Time: Nov 21 00:45:09 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N103024W.M
 Quant Title : SW846 8260
 QLast Update : Thu Oct 31 18:45:38 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.224	168	166206	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	9.100	114	309973	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.865	117	281178	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.788	152	120443	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.577	65	124313	51.785	ug/l	0.00
Spiked Amount 50.000	Range 74 - 125		Recovery = 103.560%			
35) Dibromofluoromethane	8.159	113	98461	46.928	ug/l	0.00
Spiked Amount 50.000	Range 75 - 124		Recovery = 93.860%			
50) Toluene-d8	10.565	98	362001	46.837	ug/l	0.00
Spiked Amount 50.000	Range 86 - 113		Recovery = 93.680%			
62) 4-Bromofluorobenzene	12.847	95	138588	47.979	ug/l	0.00
Spiked Amount 50.000	Range 77 - 121		Recovery = 95.960%			
Target Compounds						
16) Acetone	4.430	43	19299	25.267	ug/l	Qvalue 100
20) Methylene Chloride	5.253	84	2556m	1.211	ug/l	
43) Isopropyl Acetate	8.688	43	213072m	43.561	ug/l	

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN112024\
Data File : VN084977.D
Acq On : 20 Nov 2024 18:34
Operator : JC\MD
Sample : P4916-12
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 20 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
TP-3-WC

Manual Integrations
APPROVED

Reviewed By :Semsettin Yesilyurt 11/21/2024
Supervised By :Mahesh Dadoda 11/21/2024

Quant Time: Nov 21 00:45:09 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N103024W.M
Quant Title : SW846 8260
QLast Update : Thu Oct 31 18:45:38 2024
Response via : Initial Calibration

