

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN112024\
 Data File : VN084979.D
 Acq On : 20 Nov 2024 19:22
 Operator : JC\MD
 Sample : P4925-04
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 22 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 MH-741

Quant Time: Nov 21 00:46:24 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N103024W.M
 Quant Title : SW846 8260
 QLast Update : Thu Oct 31 18:45:38 2024
 Response via : Initial Calibration

Manual Integrations APPROVED

Reviewed By :Semsettin Yesilyurt 11/21/2024
 Supervised By :Mahesh Dadoda 11/21/2024

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.218	168	163448	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	9.100	114	307088	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.865	117	282958	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.788	152	124904	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.577	65	126255	53.481	ug/l	0.00
Spiked Amount 50.000	Range 74 - 125		Recovery =	106.960%		
35) Dibromofluoromethane	8.159	113	97960	47.128	ug/l	0.00
Spiked Amount 50.000	Range 75 - 124		Recovery =	94.260%		
50) Toluene-d8	10.565	98	366289	47.837	ug/l	0.00
Spiked Amount 50.000	Range 86 - 113		Recovery =	95.680%		
62) 4-Bromofluorobenzene	12.847	95	141159	49.328	ug/l	0.00
Spiked Amount 50.000	Range 77 - 121		Recovery =	98.660%		
Target Compounds						
16) Acetone	4.430	43	30263	42.005	ug/l	99
20) Methylene Chloride	5.265	84	2214	1.066	ug/l #	56
43) Isopropyl Acetate	8.695	43	425596m	88.828	ug/l	

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN112024\
Data File : VN084979.D
Acq On : 20 Nov 2024 19:22
Operator : JC\MD
Sample : P4925-04
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 22 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
MH-741

Manual Integrations
APPROVED

Reviewed By :Semsettin Yesilyurt 11/21/2024
Supervised By :Mahesh Dadoda 11/21/2024

Quant Time: Nov 21 00:46:24 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N103024W.M
Quant Title : SW846 8260
QLast Update : Thu Oct 31 18:45:38 2024
Response via : Initial Calibration

