

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN112024\
 Data File : VN084980.D
 Acq On : 20 Nov 2024 19:46
 Operator : JC\MD
 Sample : P4925-08
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 23 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 MH-741

Quant Time: Nov 21 00:46:53 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N103024W.M
 Quant Title : SW846 8260
 QLast Update : Thu Oct 31 18:45:38 2024
 Response via : Initial Calibration

Manual Integrations APPROVED

Reviewed By :Semsettin Yesilyurt 11/21/2024
 Supervised By :Mahesh Dadoda 11/21/2024

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.218	168	168905	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	9.100	114	307008	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.865	117	280273	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.788	152	122809	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.577	65	123662	50.690	ug/l	0.00
Spiked Amount 50.000	Range 74 - 125		Recovery = 101.380%			
35) Dibromofluoromethane	8.165	113	99557	47.908	ug/l	0.00
Spiked Amount 50.000	Range 75 - 124		Recovery = 95.820%			
50) Toluene-d8	10.565	98	365021	47.684	ug/l	0.00
Spiked Amount 50.000	Range 86 - 113		Recovery = 95.360%			
62) 4-Bromofluorobenzene	12.847	95	136874	47.843	ug/l	0.00
Spiked Amount 50.000	Range 77 - 121		Recovery = 95.680%			
Target Compounds						
16) Acetone	4.430	43	35748	48.428	ug/l #	83
20) Methylene Chloride	5.259	84	2185	1.018	ug/l #	84
43) Isopropyl Acetate	8.694	43	381042m	79.447	ug/l	

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN112024\
Data File : VN084980.D
Acq On : 20 Nov 2024 19:46
Operator : JC\MD
Sample : P4925-08
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 23 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
MH-741

Manual Integrations
APPROVED

Reviewed By :Semsettin Yesilyurt 11/21/2024
Supervised By :Mahesh Dadoda 11/21/2024

Quant Time: Nov 21 00:46:53 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N103024W.M
Quant Title : SW846 8260
QLast Update : Thu Oct 31 18:45:38 2024
Response via : Initial Calibration

