

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN112024\
 Data File : VN084981.D
 Acq On : 20 Nov 2024 20:10
 Operator : JC\MD
 Sample : P4921-01 100X
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 24 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 WC-11-A-202411

Quant Time: Nov 21 00:47:28 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N103024W.M
 Quant Title : SW846 8260
 QLast Update : Thu Oct 31 18:45:38 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.224	168	162747	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	9.100	114	302598	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.865	117	256627	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.788	152	112787	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.577	65	121372	51.634	ug/l	0.00
Spiked Amount	50.000	Range 74 - 125	Recovery	= 103.260%		
35) Dibromofluoromethane	8.165	113	97198	47.455	ug/l	0.00
Spiked Amount	50.000	Range 75 - 124	Recovery	= 94.900%		
50) Toluene-d8	10.565	98	338898	44.917	ug/l	0.00
Spiked Amount	50.000	Range 86 - 113	Recovery	= 89.840%		
62) 4-Bromofluorobenzene	12.847	95	126197	44.754	ug/l	0.00
Spiked Amount	50.000	Range 77 - 121	Recovery	= 89.500%		

Target Compounds Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN112024\
Data File : VN084981.D
Acq On : 20 Nov 2024 20:10
Operator : JC\MD
Sample : P4921-01 100X
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 24 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
WC-11-A-202411

Quant Time: Nov 21 00:47:28 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N103024W.M
Quant Title : SW846 8260
QLast Update : Thu Oct 31 18:45:38 2024
Response via : Initial Calibration

