

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN112024\
 Data File : VN084983.D
 Acq On : 20 Nov 2024 20:58
 Operator : JC\MD
 Sample : P4770-01 500X
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 26 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 MW-1

Quant Time: Nov 21 03:40:48 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N103024W.M
 Quant Title : SW846 8260
 QLast Update : Thu Oct 31 18:45:38 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.224	168	152380	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	9.100	114	277071	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.865	117	242199	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.788	152	122080	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.577	65	113952	51.776	ug/l	0.00
Spiked Amount 50.000	Range 74 - 125		Recovery = 103.560%			
35) Dibromofluoromethane	8.165	113	90226	48.109	ug/l	0.00
Spiked Amount 50.000	Range 75 - 124		Recovery = 96.220%			
50) Toluene-d8	10.565	98	319903	46.306	ug/l	0.00
Spiked Amount 50.000	Range 86 - 113		Recovery = 92.620%			
62) 4-Bromofluorobenzene	12.847	95	130351	50.486	ug/l	0.00
Spiked Amount 50.000	Range 77 - 121		Recovery = 100.980%			
Target Compounds						
						Qvalue
39) Methylcyclohexane	9.600	83	784	0.296	ug/l	93
40) Benzene	8.606	78	6541	0.783	ug/l #	68
52) Toluene	10.630	92	99603	20.389	ug/l	98
67) Ethyl Benzene	11.965	91	67442	7.457	ug/l	97
68) m/p-Xylenes	12.065	106	119638	34.939	ug/l	98
69) o-Xylene	12.394	106	52872	16.506	ug/l	100
73) Isopropylbenzene	12.694	105	21409	2.502	ug/l	98
78) n-propylbenzene	13.035	91	46927	4.681	ug/l	98
80) 1,3,5-Trimethylbenzene	13.171	105	81125	11.441	ug/l	98
84) 1,2,4-Trimethylbenzene	13.482	105	301506	41.200	ug/l	99
87) 1,3-Dichlorobenzene	13.729	146	2133	0.475	ug/l	86
91) 1,2-Dichlorobenzene	14.100	146	1977	0.458	ug/l	90
93) 1,2,4-Trichlorobenzene	15.388	180	1891	0.801	ug/l #	84
95) Naphthalene	15.635	128	59150	8.371	ug/l	98
96) 1,2,3-Trichlorobenzene	15.829	180	1905	0.831	ug/l	89

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN112024\
Data File : VN084983.D
Acq On : 20 Nov 2024 20:58
Operator : JC\MD
Sample : P4770-01 500X
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 26 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
MW-1

Quant Time: Nov 21 03:40:48 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N103024W.M
Quant Title : SW846 8260
QLast Update : Thu Oct 31 18:45:38 2024
Response via : Initial Calibration

