

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN112024\
 Data File : VN084984.D
 Acq On : 20 Nov 2024 21:22
 Operator : JC\MD
 Sample : P4892-04
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 27 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 WB-310-SW

Quant Time: Nov 21 00:51:26 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N103024W.M
 Quant Title : SW846 8260
 QLast Update : Thu Oct 31 18:45:38 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.218	168	166877	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	9.100	114	307613	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.865	117	262227	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.794	152	114908	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.577	65	127803	53.024	ug/l	0.00
Spiked Amount	50.000	Range 74 - 125	Recovery	=	106.040%	
35) Dibromofluoromethane	8.165	113	98986	47.540	ug/l	0.00
Spiked Amount	50.000	Range 75 - 124	Recovery	=	95.080%	
50) Toluene-d8	10.565	98	352034	45.897	ug/l	0.00
Spiked Amount	50.000	Range 86 - 113	Recovery	=	91.800%	
62) 4-Bromofluorobenzene	12.847	95	119438	41.666	ug/l	0.00
Spiked Amount	50.000	Range 77 - 121	Recovery	=	83.340%	
Target Compounds						
						Qvalue
68) m/p-Xylenes	12.076	106	1461	0.394	ug/l	92
84) 1,2,4-Trimethylbenzene	13.482	105	3571	0.518	ug/l	98

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN112024\
Data File : VN084984.D
Acq On : 20 Nov 2024 21:22
Operator : JC\MD
Sample : P4892-04
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 27 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
WB-310-SW

Quant Time: Nov 21 00:51:26 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N103024W.M
Quant Title : SW846 8260
QLast Update : Thu Oct 31 18:45:38 2024
Response via : Initial Calibration

