

Data Path : Z:\VOASRV\HPCHEM1\MSVOA N\DATA\VN112118\
 Data File : VN052448.D
 Acq On : 21 Nov 2018 19:37
 Operator : MD\SY
 Sample : J5763-04
 Misc : 5.00mL/MSVOA N/WATER
 ALS Vial : 24 Sample Multiplier: 28

Instrument :
 MSVOA_N
 ClientSampleId :

Quant Time: Nov 22 00:00:47 2018
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N111418W.M
 Quant Title : SW846 8260
 QLast Update : Thu Nov 15 03:02:37 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.67	168	308925	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.59	114	519439	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.41	117	439596	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.34	152	144223	50.00	ug/l	0.00

System Monitoring Compounds	R.T.	QIon	Response	Conc	Units	Dev(Min)
33) 1,2-Dichloroethane-d4	8.03	65	237035	54.33	ug/l	0.00
Spiked Amount			Recovery	=	108.66%	
35) Dibromofluoromethane	7.59	113	171861	51.30	ug/l	0.00
Spiked Amount			Recovery	=	102.60%	
50) Toluene-d8	10.09	98	657337	49.88	ug/l	0.00
Spiked Amount			Recovery	=	99.76%	
62) 4-Bromofluorobenzene	12.40	95	190599	43.96	ug/l	0.00
Spiked Amount			Recovery	=	87.92%	


Target Compounds	R.T.	QIon	Response	Conc	Units	Qvalue
13) Acrolein	3.60	56	29	21.847	ug/l #	14
16) Acetone	3.82	43	5641	4.696	ug/l	99
18) Methyl Acetate	4.33	43	1866	0.981	ug/l #	64
20) Methylene Chloride	4.56	84	76132	20.416	ug/l	94
29) Tetrahydrofuran	7.21	42	350	0.315	ug/l #	40
31) Cyclohexane	7.67	56	10207	0.489	ug/l #	53
36) 1,1-Dichloropropene	7.67	75	29251	5.618	ug/l #	51
38) Carbon Tetrachloride	7.67	117	34814	7.278	ug/l #	16
41) Methacrylonitrile	7.19	41	131	3.753	ug/l #	100
43) Isopropyl Acetate	8.17	43	17630	2.418	ug/l #	96
48) Methyl methacrylate	9.18	41	5681	1.536	ug/l #	16
51) 4-Methyl-2-Pentanone	10.02	43	5523	1.505	ug/l #	41
52) Toluene	10.15	92	3558	0.400	ug/l	94
54) cis-1,3-Dichloropropene	9.90	75	31931	5.539	ug/l #	76
57) 1,3-Dichloropropane	10.76	76	6883	1.205	ug/l #	43
59) 2-Hexanone	10.76	43	123188	48.798	ug/l #	55
68) m/p-Xylenes	11.62	106	2093	0.355	ug/l	99
69) o-Xylene	11.94	106	1243	0.220	ug/l	99
74) N-amyl acetate	11.88	43	14142	3.424	ug/l #	58
76) 1,2,3-Trichloropropane	12.40	75	110566	45.181	ug/l #	100
80) 1,3,5-Trimethylbenzene	12.73	105	4857	0.527	ug/l	99
81) trans-1,4-Dichloro-2-buten	12.40	75	110566	146.539	ug/l #	5
84) 1,2,4-Trimethylbenzene	13.04	105	8204	0.886	ug/l	97
95) Naphthalene	15.13	128	3157	3.191	ug/l	97


(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA N\DATA\VN112118\
Data File : VN052448.D
Acq On : 21 Nov 2018 19:37
Operator : MD\SY
Sample : J5763-04
Misc : 5.00mL/MSVOA N/WATER
ALS Vial : 24 Sample Multiplier: 28

Instrument :
MSVOA_N
ClientSampleId :

Quant Time: Nov 22 00:00:47 2018
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_N\METHODS\82N111418W.M
Quant Title : SW846 8260
QLast Update : Thu Nov 15 03:02:37 2018
Response via : Initial Calibration





#1

Pentafluorobenzene

Concen: 50.000 ug/l

RT: 7.67 min Scan# 1833

Delta R.T. 0.00 min

Lab File: VN052448.D

Acq: 21 Nov 2018 19:37

Instrument :

MSVOA_N


ClientSampleId :

Tot Ion:168 Resp: 308925

Ion Ratio Lower Upper




168 100

99 64.0 47.9 71.9



Abundance Ion 167.90 (167.60 to 168.60): V

Ion 98.90 (98.60 to 99.60): VN0



#13

Acrolein

Concen: 21.847 ug/l

RT: 3.60 min Scan# 569

Delta R.T. -0.01 min

Lab File: VN052448.D


Acq: 21 Nov 2018 19:37

Tgt Ion: 56 Resp: 29

Ion Ratio Lower Upper



56 100


55 0.0 57.6 86.4#



Abundance Ion 56.00 (55.70 to 56.70): VN0

Ion 55.00 (54.70 to 55.70): VN0





#16

Acetone

Concen: 4.696 ug/l

RT: 3.82 min Scan# 635

Delta R.T. 0.00 min

Lab File: VN052448.D

Acq: 21 Nov 2018 19:37

Instrument :

MSVOA_N





ClientSampled :

Tot Ion: 43 Resp: 5641

Ion Ratio Lower Upper

43 100

58 31.1 24.2 36.4



#18

Methyl Acetate

Concen: 0.981 ug/l

RT: 4.33 min Scan# 796

Delta R.T. 0.01 min

Lab File: VN052448.D




Acq: 21 Nov 2018 19:37


Tgt Ion: 43 Resp: 1866

Ion Ratio Lower Upper

43 100

74 3.4 15.9 23.9#





#20

Methylene Chloride

Concen: 20.416 ug/l

RT: 4.56 min Scan# 866

Delta R.T. 0.00 min

Lab File: VN052448.D

Acq: 21 Nov 2018 19:37

Instrument :
MSVOA_N
ClientSampled :

Tot Ion: 84 Resp: 76132


Ion Ratio Lower Upper

84 100

49 157.8 118.2 177.2

51 46.2 36.1 54.1

86 60.3 51.6 77.4






Abundance Ion 83.90 (83.60 to 84.60): VN0

Ion 48.90 (48.60 to 49.60): VN0

Ion 50.90 (50.60 to 51.60): VN0

Ion 85.90 (85.60 to 86.60): VN0



#29

Tetrahydrofuran

Concen: 0.315 ug/l

RT: 7.21 min Scan# 1690

Delta R.T. -0.00 min

Lab File: VN052448.D

Acq: 21 Nov 2018 19:37


Tgt Ion: 42 Resp: 350

Ion Ratio Lower Upper

42 100

72 0.0 28.6 42.8#



71 0.0 26.7 40.1#




Abundance Ion 42.00 (41.70 to 42.70): VN0

Ion 72.00 (71.70 to 72.70): VN0

Ion 71.00 (70.70 to 71.70): VN0









#31
Cyclohexane
Concen: 0.489 ug/l
RT: 7.67 min Scan# 1833
Delta R.T. 0.01 min
Lab File: VN052448.D
Acq: 21 Nov 2018 19:37

Instrument :
MSVOA_N
ClientSampleId :

Tot Ion: 56 Resp: 10207
Ion Ratio Lower Upper
56 100
69 113.9 23.8 35.8#
84 84.4 63.0 94.6




Abundance Ion 56.00 (55.70 to 56.70): VN0
Ion 69.00 (68.70 to 69.70): VN0
Ion 84.00 (83.70 to 84.70): VN0






#33
1,2-Dichloroethane-d4
Concen: Below ug/l
RT: 8.03 min Scan# 1945
Delta R.T. 0.00 min
Lab File: VN052448.D
Acq: 21 Nov 2018 19:37

Tgt Ion: 65 Resp: 237035
Ion Ratio Lower Upper
65 100
67 50.4 0.0 105.2



Abundance Ion 64.90 (64.60 to 65.60): VN0
Ion 66.90 (66.60 to 67.60): VN0





#34

1,4-Difluorobenzene

Concen: 50.000 ug/l

RT: 8.59 min Scan# 2119

Delta R.T. 0.00 min

Lab File: VN052448.D

Acq: 21 Nov 2018 19:37

Instrument :

MSVOA_N

ClientSampleId :


Tot Ion:114 Resp: 519439

Ion Ratio Lower Upper

114 100

63 23.1 0.0 45.4




88 16.2 0.0 31.2



Abundance Ion 113.90 (113.60 to 114.60): V

Ion 63.00 (62.70 to 63.70): VN

Ion 87.90 (87.60 to 88.60): VN



#35

Dibromofluoromethane

Concen: 50.000 ug/l

RT: 7.59 min Scan# 1810

Delta R.T. 0.00 min

Lab File: VN052448.D

Acq: 21 Nov 2018 19:37


Tgt Ion:113 Resp: 171861

Ion Ratio Lower Upper

113 100

111 105.1 81.0 121.6



192 17.4 13.8 20.6




Abundance Ion 112.80 (112.50 to 113.50): V

Ion 110.80 (110.50 to 111.50): V

Ion 191.70 (191.40 to 192.40): V









#36
 1,1-Dichloropropene
 Concen: 5.618 ug/l
 RT: 7.67 min Scan# 1833
 Delta R.T. -0.13 min
 Lab File: VN052448.D
 Acq: 21 Nov 2018 19:37

Instrument :
 MSVOA_N
 ClientSampleId :

Tot Ion	Ratio	Lower	Upper
75	100		
110	8.6	16.5	49.5#
77	0.0	24.2	36.2#




Abundance Ion 74.90 (74.60 to 75.60): VN052307.D (-1849) (-)
 Ion 109.90 (109.60 to 110.60): VN052307.D (-1849) (-)
 Ion 76.90 (76.60 to 77.60): VN052307.D (-1849) (-)






#38
 Carbon Tetrachloride
 Concen: 7.278 ug/l
 RT: 7.67 min Scan# 1833
 Delta R.T. -0.11 min
 Lab File: VN052448.D
 Acq: 21 Nov 2018 19:37

Tgt Ion	Ratio	Lower	Upper
117	100		
119	5.3	77.4	116.0#
121	0.0	24.0	36.0#



Abundance Ion 116.80 (116.50 to 117.50): VN052307.D (-1848) (-)
 Ion 118.80 (118.50 to 119.50): VN052307.D (-1848) (-)
 Ion 120.80 (120.50 to 121.50): VN052307.D (-1848) (-)





#41

Methacrylonitrile

Concen: 3.753 ug/l

RT: 7.19 min Scan# 1685

Delta R.T. 0.02 min

Lab File: VN052448.D

Acq: 21 Nov 2018 19:37

Instrument :
MSVOA_N
ClientSampled :

Tot Ion: 41 Resp: 131





Ion Ratio Lower Upper

41 100

39 0.0 0.0 0.0

67 0.0 0.0 0.0

52 0.0 0.0 0.0



#43

Isopropyl Acetate

Concen: 2.418 ug/l

RT: 8.17 min Scan# 1990

Delta R.T. 0.01 min

Lab File: VN052448.D

Acq: 21 Nov 2018 19:37




Tgt Ion: 43 Resp: 17630


Ion Ratio Lower Upper

43 100

61 16.4 12.0 18.0

87 8.2 8.3 12.5#









#48
Methvl methacrylate
Concen: 1.536 ug/l
RT: 9.18 min Scan# 2305
Delta R.T. -0.01 min
Lab File: VN052448.D
Acq: 21 Nov 2018 19:37

Instrument :
MSVOA_N
ClientSampleId :

Tot Ion: 41 Resp: 5681
Ion Ratio Lower Upper
41 100
69 0.0 57.1 85.7#
39 0.0 52.5 78.7#




Abundance Ion 41.00 (40.70 to 41.70): VN0
Ion 69.00 (68.70 to 69.70): VN0
Ion 39.00 (38.70 to 39.70): VN0






#50
Toluene-d8
Concen: N.D. ug/l
RT: 10.09 min Scan# 2587
Delta R.T. 0.00 min
Lab File: VN052448.D
Acq: 21 Nov 2018 19:37

Tgt Ion: 98 Resp: 657337
Ion Ratio Lower Upper
98 100
100 62.3 49.5 74.3



Abundance Ion 98.00 (97.70 to 98.70): VN0
Ion 100.00 (99.70 to 100.70): VN0





#51

4-Methyl-2-Pentanone

Concen: 1.505 ug/l

RT: 10.02 min Scan# 2565

Delta R.T. 0.04 min

Lab File: VN052448.D

Acq: 21 Nov 2018 19:37

Instrument :

MSVOA_N


ClientSampled :

Tot Ion: 43 Resp: 5523

Ion Ratio Lower Upper




43 100

58 0.0 27.0 40.6#



Abundance Ion 43.00 (42.70 to 43.70): VN0

Ion 58.00 (57.70 to 58.70): VN0



#52

Toluene

Concen: 0.400 ug/l

RT: 10.15 min Scan# 2606

Delta R.T. -0.00 min

Lab File: VN052448.D


Acq: 21 Nov 2018 19:37

Tgt Ion: 92 Resp: 3558

Ion Ratio Lower Upper



92 100


91 170.3 143.6 215.4



Abundance Ion 92.00 (91.70 to 92.70): VN0

Ion 91.00 (90.70 to 91.70): VN0






#54




cis-1,3-Dichloropropene
 Concen: 5.539 ug/l
 RT: 9.90 min Scan# 2528
 Delta R.T. 0.06 min
 Lab File: VN052448.D
 Acq: 21 Nov 2018 19:37

Instrument :
 MSVOA_N
 ClientSampled :

Tot Ion: 75 Resp: 31931
 Ion Ratio Lower Upper
 75 100
 77 0.0 24.6 36.8#
 39 56.6 49.9 74.9




Abundance Ion 74.90 (74.60 to 75.60): VN0
 Ion 76.90 (76.60 to 77.60): VN0
 Ion 39.00 (38.70 to 39.70): VN0





#57


1,3-Dichloropropane
 Concen: 1.205 ug/l
 RT: 10.76 min Scan# 2795
 Delta R.T. 0.05 min
 Lab File: VN052448.D
 Acq: 21 Nov 2018 19:37

Tgt Ion: 76 Resp: 6883
 Ion Ratio Lower Upper
 76 100
 78 0.0 25.7 38.5#



Abundance Ion 75.90 (75.60 to 76.60): VN0
 Ion 77.90 (77.60 to 78.60): VN0









#59
2-Hexanone
Concen: 48.798 ug/l
RT: 10.76 min Scan# 2796
Delta R.T. 0.01 min
Lab File: VN052448.D
Acq: 21 Nov 2018 19:37




Instrument :
MSVOA_N
ClientSampled :


Tot Ion: 43 Resp: 123188
Ion Ratio Lower Upper
43 100
58 17.0 23.5 70.6#



#62
4-Bromofluorobenzene
Concen: N.D. ug/l
RT: 12.40 min Scan# 3305
Delta R.T. 0.00 min
Lab File: VN052448.D
Acq: 21 Nov 2018 19:37

Tgt Ion: 95 Resp: 190599
Ion Ratio Lower Upper
95 100
174 74.1 0.0 156.0
176 70.3 0.0 148.4





#63

Chlorobenzene-d5

Concen: 50.000 ug/l

RT: 11.41 min Scan# 2996

Delta R.T. 0.00 min

Lab File: VN052448.D

Acq: 21 Nov 2018 19:37

Instrument :

MSVOA_N

ClientSampleId :


Tot Ion:117 Resp: 439596

Ion Ratio Lower Upper

117 100

82 60.3 46.5 69.7




119 31.5 25.7 38.5



Abundance Ion 116.90 (116.60 to 117.60): V

Ion 82.00 (81.70 to 82.70): VN

Ion 118.90 (118.60 to 119.60): V



#68

m/p-Xylenes

Concen: 0.355 ug/l

RT: 11.62 min Scan# 3061

Delta R.T. -0.00 min

Lab File: VN052448.D


Acq: 21 Nov 2018 19:37

Tgt Ion:106 Resp: 2093

Ion Ratio Lower Upper



106 100


91 208.0 165.5 248.3



Abundance Ion 106.00 (105.70 to 106.70): V

Ion 91.00 (90.70 to 91.70): VN









#69
o-Xylene
Concen: 0.220 ug/l
RT: 11.94 min Scan# 3163
Delta R.T. -0.00 min
Lab File: VN052448.D
Acq: 21 Nov 2018 19:37

Instrument :
MSVOA_N
ClientSampled :

Tot Ion:106 Resp: 1243
Ion Ratio Lower Upper
106 100
91 222.4 110.0 330.0




Abundance Ion 106.00 (105.70 to 106.70): V
Ion 91.00 (90.70 to 91.70): V






#72
1,4-Dichlorobenzene-d4
Concen: 50.000 ug/l
RT: 13.34 min Scan# 3598
Delta R.T. 0.00 min
Lab File: VN052448.D
Acq: 21 Nov 2018 19:37

Tgt Ion:152 Resp: 144223
Ion Ratio Lower Upper
152 100
115 61.5 30.3 91.0
150 157.3 0.0 348.2



Abundance Ion 151.90 (151.60 to 152.60): V
Ion 114.90 (114.60 to 115.60): V
Ion 149.90 (149.60 to 150.60): V





#74

N-amvl acetate

Concen: 3.424 ug/l

RT: 11.88 min Scan# 3142

Delta R.T. -0.19 min

Lab File: VN052448.D

Acq: 21 Nov 2018 19:37

Instrument :

MSVOA_N

ClientSampled :

Tot Ion: 43 Resp: 14142


Ion Ratio Lower Upper

43 100

70 1.8 27.1 40.7#

55 5.7 17.4 26.0#

61 8.4 16.1 24.1#






Abundance Ion 43.00 (42.70 to 43.70): VN0

Ion 70.00 (69.70 to 70.70): VN0

Ion 55.00 (54.70 to 55.70): VN0

Ion 61.00 (60.70 to 61.70): VN0



#76

1,2,3-Trichloropropane

Concen: 45.181 ug/l

RT: 12.40 min Scan# 3305

Delta R.T. -0.15 min

Lab File: VN052448.D


Acq: 21 Nov 2018 19:37

Tgt Ion: 75 Resp: 110566

Ion Ratio Lower Upper



75 100


77 1.2 0.0 0.0#



Abundance Ion 74.90 (74.60 to 75.60): VN0

Ion 77.00 (76.70 to 77.70): VN0





#80

1,3,5-Trimethylbenzene

Concen: 0.527 ug/l

RT: 12.73 min Scan# 3407

Delta R.T. -0.00 min

Lab File: VN052448.D

Acq: 21 Nov 2018 19:37

Instrument :

MSVOA_N


ClientSampleId :

Tot Ion:105 Resp: 4857

Ion Ratio Lower Upper




105 100

120 48.3 23.7 71.1



Abundance Ion 105.00 (104.70 to 105.70): V

Ion 120.00 (119.70 to 120.70): V



#81

trans-1,4-Dichloro-2-butene

Concen: 146.539 ug/l

RT: 12.40 min Scan# 3305

Delta R.T. 0.10 min

Lab File: VN052448.D

Acq: 21 Nov 2018 19:37


Tgt Ion: 75 Resp: 110566

Ion Ratio Lower Upper

75 100

53 0.1 90.2 135.2#



89 0.0 35.4 53.0#




Abundance Ion 74.90 (74.60 to 75.60): VNO

Ion 53.00 (52.70 to 53.70): VNO

Ion 88.90 (88.60 to 89.60): VNO









#84
 1,2,4-Trimethylbenzene
 Concen: 0.886 ug/l
 RT: 13.04 min Scan# 3504
 Delta R.T. 0.00 min
 Lab File: VN052448.D
 Acq: 21 Nov 2018 19:37

Instrument :
 MSVOA_N
 ClientSampled :

Tot Ion:105 Resp: 8204
 Ion Ratio Lower Upper
 105 100
 120 42.9 22.5 67.5



#95
 Naphthalene
 Concen: 3.191 ug/l
 RT: 15.13 min Scan# 4155
 Delta R.T. 0.00 min
 Lab File: VN052448.D
 Acq: 21 Nov 2018 19:37

Tgt Ion:128 Resp: 3157
 Ion Ratio Lower Upper
 128 100
 127 11.0 10.4 15.6
 129 11.2 8.6 13.0

