

Data Path : Z:\VOASRV\HPCHEM1\MSVOA_N\DATA\VN112118\
 Data File : VN052432.D
 Acq On : 21 Nov 2018 12:31
 Operator : MD\SY
 Sample : PB114992ZHE#02
 Misc : 5.00mL/MSVOA_N/WATER
 ALS Vial : 8 Sample Multiplier: 28

Instrument :
 MSVOA_N
 ClientSampleId :
 PB114992ZHE#02

Quant Time: Nov 22 00:27:35 2018
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N111418W.M
 Quant Title : SW846 8260
 QLast Update : Thu Nov 15 03:02:37 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.67	168	312569	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.59	114	529162	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.41	117	457728	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.35	152	156789	50.00	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.03	65	237821	53.87	ug/l	0.00
Spiked Amount	50.000		Recovery	=	107.74%	
35) Dibromofluoromethane	7.59	113	174091	51.01	ug/l	0.00
Spiked Amount	50.000		Recovery	=	102.02%	
50) Toluene-d8	10.09	98	670868	49.97	ug/l	0.00
Spiked Amount	50.000		Recovery	=	99.94%	
62) 4-Bromofluorobenzene	12.40	95	200305	45.35	ug/l	0.00
Spiked Amount	50.000		Recovery	=	90.70%	

Target Compounds						Qvalue
20) Methylene Chloride	4.55	84	11784	2.52	ug/l	89
43) Isopropyl Acetate	8.17	43	18276	2.46	ug/l	# 97

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA_N\DATA\VN112118\
Data File : VN052432.D
Acq On : 21 Nov 2018 12:31
Operator : MD\SY
Sample : PB114992ZHE#02
Misc : 5.00mL/MSVOA_N/WATER
ALS Vial : 8 Sample Multiplier: 28

Instrument :
MSVOA_N
ClientSampleId :
PB114992ZHE#02

Quant Time: Nov 22 00:27:35 2018
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_N\METHODS\82N111418W.M
Quant Title : SW846 8260
QLast Update : Thu Nov 15 03:02:37 2018
Response via : Initial Calibration

