

Data Path : Z:\VOASRV\HPCHEM1\MSVOA_N\DATA\VN112118\
 Data File : VN052433.D
 Acq On : 21 Nov 2018 12:57
 Operator : MD\SY
 Sample : PB114992ZHE#03
 Misc : 5.00mL/MSVOA_N/WATER
 ALS Vial : 9 Sample Multiplier: 28

Instrument :
 MSVOA_N
 ClientSampleId :
 PB114992ZHE#03

Quant Time: Nov 22 00:29:15 2018
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N111418W.M
 Quant Title : SW846 8260
 QLast Update : Thu Nov 15 03:02:37 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.67	168	317386	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.59	114	534824	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.41	117	461975	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.35	152	159560	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.03	65	246001	54.88	ug/l	0.00
Spiked Amount 50.000			Recovery	=	109.76%	
35) Dibromofluoromethane	7.59	113	177453	51.44	ug/l	0.00
Spiked Amount 50.000			Recovery	=	102.88%	
50) Toluene-d8	10.09	98	679448	50.07	ug/l	0.00
Spiked Amount 50.000			Recovery	=	100.14%	
62) 4-Bromofluorobenzene	12.40	95	204402	45.79	ug/l	0.00
Spiked Amount 50.000			Recovery	=	91.58%	
Target Compounds						
20) Methylene Chloride	4.55	84	11900	2.50	ug/l	90
43) Isopropyl Acetate	8.17	43	19467	2.59	ug/l #	90

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA_N\DATA\VN112118\
Data File : VN052433.D
Acq On : 21 Nov 2018 12:57
Operator : MD\SY
Sample : PB114992ZHE#03
Misc : 5.00mL/MSVOA_N/WATER
ALS Vial : 9 Sample Multiplier: 28

Instrument :
MSVOA_N
ClientSampleId :
PB114992ZHE#03

Quant Time: Nov 22 00:29:15 2018
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_N\METHODS\82N111418W.M
Quant Title : SW846 8260
QLast Update : Thu Nov 15 03:02:37 2018
Response via : Initial Calibration

