

Data Path : Z:\VOASRV\HPCHEM1\MSVOA_N\DATA\VN112118\
 Data File : VN052434.D
 Acq On : 21 Nov 2018 13:24
 Operator : MD\SY
 Sample : PB114992ZHE#04
 Misc : 5.00mL/MSVOA_N/WATER
 ALS Vial : 10 Sample Multiplier: 28

Instrument :
 MSVOA_N
 ClientSampleId :
 PB114992ZHE#04

Quant Time: Nov 22 00:31:19 2018
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N111418W.M
 Quant Title : SW846 8260
 QLast Update : Thu Nov 15 03:02:37 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.67	168	315897	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.59	114	526927	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.41	117	456996	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.35	152	153789	50.00	ug/l	0.00

System Monitoring Compounds	R.T.	QIon	Response	Conc	Units	Dev(Min)
33) 1,2-Dichloroethane-d4	8.03	65	241527	54.14	ug/l	0.00
Spiked Amount 50.000			Recovery	=	108.28%	
35) Dibromofluoromethane	7.59	113	177062	52.10	ug/l	0.00
Spiked Amount 50.000			Recovery	=	104.20%	
50) Toluene-d8	10.09	98	674774	50.47	ug/l	0.00
Spiked Amount 50.000			Recovery	=	100.94%	
62) 4-Bromofluorobenzene	12.40	95	198667	45.17	ug/l	0.00
Spiked Amount 50.000			Recovery	=	90.34%	

Target Compounds	R.T.	QIon	Response	Conc	Units	Qvalue
20) Methylene Chloride	4.56	84	11183	2.32	ug/l	95
43) Isopropyl Acetate	8.17	43	17929	2.42	ug/l	# 96

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA_N\DATA\VN112118\
Data File : VN052434.D
Acq On : 21 Nov 2018 13:24
Operator : MD\SY
Sample : PB114992ZHE#04
Misc : 5.00mL/MSVOA_N/WATER
ALS Vial : 10 Sample Multiplier: 28

Instrument :
MSVOA_N
ClientSampleId :
PB114992ZHE#04

Quant Time: Nov 22 00:31:19 2018
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_N\METHODS\82N111418W.M
Quant Title : SW846 8260
QLast Update : Thu Nov 15 03:02:37 2018
Response via : Initial Calibration

