

Data Path : Z:\VOASRV\HPCHEM1\MSVOA_N\DATA\VN112118\
 Data File : VN052443.D
 Acq On : 21 Nov 2018 17:24
 Operator : MD\SY
 Sample : J5768-06
 Misc : 5.00mL/MSVOA_N/WATER
 ALS Vial : 19 Sample Multiplier: 28

Instrument :
 MSVOA_N
 ClientSampleId :
 OK-03-112018

Quant Time: Nov 22 01:06:26 2018
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N111418W.M
 Quant Title : SW846 8260
 QLast Update : Thu Nov 15 03:02:37 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.67	168	313703	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.59	114	530473	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.41	117	456114	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.34	152	152038	50.00	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.03	65	236867	53.46	ug/l	0.00
Spiked Amount	50.000		Recovery	=	106.92%	
35) Dibromofluoromethane	7.59	113	174694	51.06	ug/l	0.00
Spiked Amount	50.000		Recovery	=	102.12%	
50) Toluene-d8	10.09	98	668636	49.68	ug/l	0.00
Spiked Amount	50.000		Recovery	=	99.36%	
62) 4-Bromofluorobenzene	12.40	95	196848	44.46	ug/l	0.00
Spiked Amount	50.000		Recovery	=	88.92%	

Target Compounds						Qvalue
16) Acetone	3.82	43	9395	7.70	ug/l #	83
20) Methylene Chloride	4.55	84	21566596	5892.91	ug/l	87
43) Isopropyl Acetate	8.17	43	22440	3.01	ug/l #	91
95) Naphthalene	15.14	128	5794	3.63	ug/l #	91

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA_N\DATA\VN112118\
Data File : VN052443.D
Acq On : 21 Nov 2018 17:24
Operator : MD\SY
Sample : J5768-06
Misc : 5.00mL/MSVOA_N/WATER
ALS Vial : 19 Sample Multiplier: 28

Instrument :
MSVOA_N
ClientSampleId :
OK-03-112018

Quant Time: Nov 22 01:06:26 2018
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_N\METHODS\82N111418W.M
Quant Title : SW846 8260
QLast Update : Thu Nov 15 03:02:37 2018
Response via : Initial Calibration

