

Data Path : Z:\VOASRV\HPCHEM1\MSVOA_N\DATA\VN112118\
 Data File : VN052449.D
 Acq On : 21 Nov 2018 20:03
 Operator : MD\SY
 Sample : J5996-05 5X
 Misc : 5.00mL/MSVOA_N/WATER
 ALS Vial : 25 Sample Multiplier: 28

Instrument :
 MSVOA_N
 ClientSampleId :
 AOC2-01R

Manual Integrations
 APPROVED

MMDadoda
 11/26/2018 4:56:45 PM

Quant Time: Nov 22 01:25:42 2018
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N111418W.M
 Quant Title : SW846 8260
 QLast Update : Thu Nov 15 03:02:37 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.67	168	328935	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.59	114	543756	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.41	117	474534	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.34	152	200301	50.00	ug/l	0.00

System Monitoring Compounds	R.T.	QIon	Response	Conc	Units	Dev(Min)
33) 1,2-Dichloroethane-d4	8.03	65	239248	51.50	ug/l	0.00
Spiked Amount			Recovery	=	103.00%	
35) Dibromofluoromethane	7.59	113	175339	49.99	ug/l	0.00
Spiked Amount			Recovery	=	99.98%	
50) Toluene-d8	10.09	98	694306	50.32	ug/l	0.00
Spiked Amount			Recovery	=	100.64%	
62) 4-Bromofluorobenzene	12.40	95	230270	50.73	ug/l	0.00
Spiked Amount			Recovery	=	101.46%	

Target Compounds	R.T.	QIon	Response	Conc	Units	Qvalue
16) Acetone	3.81	43	8883	6.94	ug/l #	65
20) Methylene Chloride	4.55	84	33922	8.13	ug/l	98
39) Methylcyclohexane	9.08	83	17978	2.89	ug/l #	81
40) Benzene	8.04	78	2680550	172.16	ug/l	98
42) 1,2-Dichloroethane	8.13	62	18003	3.24	ug/l	92
52) Toluene	10.16	92	7103050	762.62	ug/l	98
61) 1,2-Dibromoethane	11.01	107	42851	12.74	ug/l	96
67) Ethyl Benzene	11.51	91	1253245	71.99	ug/l	100
68) m/p-Xylenes	11.62	106	1401260	220.29	ug/l	96
69) o-Xylene	11.95	106	687474	112.75	ug/l	96
73) Isopropylbenzene	12.25	105	57116	3.55	ug/l	100
78) n-propylbenzene	12.59	91	185508	9.85	ug/l	98
80) 1,3,5-Trimethylbenzene	12.73	105	244724	19.11	ug/l	100
84) 1,2,4-Trimethylbenzene	13.04	105	899581	69.97	ug/l	98
85) sec-Butylbenzene	13.17	105	7261m	0.47	ug/l	
95) Naphthalene	15.13	128	132155	20.27	ug/l	99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA_N\DATA\VN112118\
Data File : VN052449.D
Acq On : 21 Nov 2018 20:03
Operator : MD\SY
Sample : J5996-05 5X
Misc : 5.00mL/MSVOA_N/WATER
ALS Vial : 25 Sample Multiplier: 28

Instrument :
MSVOA_N
ClientSampleId :
AOC2-01R

Manual Integrations
APPROVED

MMDadoda
11/26/2018 4:56:45 PM

Quant Time: Nov 22 01:25:42 2018
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_N\METHODS\82N111418W.M
Quant Title : SW846 8260
QLast Update : Thu Nov 15 03:02:37 2018
Response via : Initial Calibration

