

Data Path : Z:\VOASRV\HPCHEM1\MSVOA_N\DATA\VN112118\
 Data File : VN052450.D
 Acq On : 21 Nov 2018 20:30
 Operator : MD\SY
 Sample : J5996-13 5X
 Misc : 5.00mL/MSVOA_N/WATER
 ALS Vial : 26 Sample Multiplier: 28

Instrument :
 MSVOA_N
 ClientSampleId :
 AOC1-05-A

Quant Time: Nov 22 01:27:42 2018
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N111418W.M
 Quant Title : SW846 8260
 QLast Update : Thu Nov 15 03:02:37 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.67	168	324235	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.59	114	544552	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.41	117	460556	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.34	152	158225	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.03	65	240894	52.60	ug/l	0.00
Spiked Amount 50.000			Recovery	=	105.20%	
35) Dibromofluoromethane	7.59	113	178181	50.73	ug/l	0.00
Spiked Amount 50.000			Recovery	=	101.46%	
50) Toluene-d8	10.09	98	678552	49.11	ug/l	0.00
Spiked Amount 50.000			Recovery	=	98.22%	
62) 4-Bromofluorobenzene	12.40	95	199055	43.79	ug/l	0.00
Spiked Amount 50.000			Recovery	=	87.58%	
Target Compounds						
20) Methylene Chloride	4.56	84	21841	5.064	ug/l	98
52) Toluene	10.16	92	14052	1.506	ug/l	87

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA_N\DATA\VN112118\
Data File : VN052450.D
Acq On : 21 Nov 2018 20:30
Operator : MD\SY
Sample : J5996-13 5X
Misc : 5.00mL/MSVOA_N/WATER
ALS Vial : 26 Sample Multiplier: 28

Instrument :
MSVOA_N
ClientSampleId :
AOC1-05-A

Quant Time: Nov 22 01:27:42 2018
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_N\METHODS\82N111418W.M
Quant Title : SW846 8260
QLast Update : Thu Nov 15 03:02:37 2018
Response via : Initial Calibration

