

Data Path : Z:\VOASRV\HPCHEM1\MSVOA_N\DATA\VN112119\
 Data File : VN059330.D
 Acq On : 21 Nov 2019 13:06
 Operator : JC/SP
 Sample : VN1121MBL01
 Misc : 5.00g/10mL/100uL/5.00mL/MSVOA_N/MEOH
 ALS Vial : 4 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 VN1121MBL01

Quant Time: Nov 22 06:19:05 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N112019W.M
 Quant Title : SW846 8260
 QLast Update : Thu Nov 21 08:44:45 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.64	168	410301	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.57	114	636123	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.40	117	546244	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.35	152	199814	50.00	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.01	65	237798	48.30	ug/l	0.00
Spiked Amount	50.000		Recovery	=	96.60%	
35) Dibromofluoromethane	7.57	113	187663	49.17	ug/l	0.00
Spiked Amount	50.000		Recovery	=	98.34%	
50) Toluene-d8	10.08	98	767870	50.61	ug/l	0.00
Spiked Amount	50.000		Recovery	=	101.22%	
62) 4-Bromofluorobenzene	12.40	95	245979	46.37	ug/l	0.00
Spiked Amount	50.000		Recovery	=	92.74%	

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA N\DATA\VN112119\
Data File : VN059330.D
Acq On : 21 Nov 2019 13:06
Operator : JC/SP
Sample : VN1121MBL01
Misc : 5.00g/10mL/100uL/5.00mL/MSVOA_N/MEOH
ALS Vial : 4 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
VN1121MBL01

Quant Time: Nov 22 06:19:05 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_N\METHODS\82N112019W.M
Quant Title : SW846 8260
QLast Update : Thu Nov 21 08:44:45 2019
Response via : Initial Calibration

